

LAB MONTHLY SUMMARY

Lab Name **JEEVAN DIAGNOSTIC CENTRE**Lab No **13184**Month **March**Year **2020**Constituent Group **Chemistry II**Date of Result Entered : **16/03/2020**Date of Report Published : **11/04/2020**

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE I	GOD-POD	Any Analyser	950	284.86	12.45	35.47	278 mg/dl	-0.25	2.30
2	UREA I	UREASE UV / GLDH	Any Analyser	552	60.76	12.51	7.60	57.6 mg/dl	-0.42	0.65
3	CREATININE I	OTHERS (Any Other Principles / Methods)	Any Analyser	126	2.81	14.43	0.41	2.8 mg/dl	-0.02	0.07
4	T-PROTEIN I	BIURET - colorimetric	Any Analyser	723	5.11	11.65	0.60	4.9 g/dl	-0.35	0.04
5	ALBUMIN I	BCG - colorimetric	Any Analyser	705	3.09	10.42	0.32	2.9 g/dl	-0.59	0.02
6	URIC ACID I	ENZYMATIC / URICASE Colorimetric	Any Analyser	660	4.16	23.53	0.98	4.2 mg/dl	0.04	0.08
7	CHOLESTEROL I	CHOD-PAP	Any Analyser	835	106.37	12.04	12.80	103.6 mg/dl	-0.22	0.89
8	TRIGLYCERIDE I	GPO-PAP / Enzymatic Colorimetric / End Point	Any Analyser	731	99.19	14.13	14.02	82 mg/dl	-1.23	1.04

SDI Range	Interpretation
Within -1.0 to +1.0	Excellent.
Between ± 1.0 to ± 2.0	Good.
Between ± 2.0 to ± 3.0	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

Page 1 of 1

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

Contact details:

Email: clinqc@cmcvellore.ac.in

Contact Number: 0416-2283102

Pamela Christudoss

Dr. Pamela Christudoss
 CMC EQAS Co-Ordinator
 Christian Medical College, Vellore

***** End of Report *****