

LAB MONTHLY SUMMARY

Lab Name **JEEVAN DIAGNOSTIC CENTRE**Lab No **13184**Month **April**Year **2020**Constituent Group **Chemistry II**Date of Result Entered : **21/05/2020**Date of Report Published : **06/06/2020**

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE I	GOD-POD	Any Analyser	855	190.13	11.09	21.09	189.6 mg/dl	-0.03	1.44
2	UREA I	UREASE UV / GLDH	Any Analyser	544	75.79	14.56	11.04	82.5 mg/dl	0.61	0.95
3	CREATININE I	OTHERS (Any Other Principles / Methods)	Any Analyser	114	1.36	15.38	0.21	1.38 mg/dl	0.10	0.04
4	T-PROTEIN I	BIURET - colorimetric	Any Analyser	672	4.97	12.19	0.61	4.56 g/dl	-0.68	0.05
5	ALBUMIN I	BCG - colorimetric	Any Analyser	626	2.99	11.11	0.33	2.6 g/dl	-1.17	0.03
6	URIC ACID I	ENZYMATIC / URICASE Colorimetric	Any Analyser	640	7.92	14.99	1.19	6.8 mg/dl	-0.94	0.09
7	CHOLESTEROL I	CHOD-PAP	Any Analyser	752	95.23	15.16	14.44	102.6 mg/dl	0.51	1.05
8	TRIGLYCERIDE I	GPO-PAP / Enzymatic Colorimetric / End Point	Any Analyser	706	175.65	13.50	23.72	189.6 mg/dl	0.59	1.79

SDI Range	Interpretation
Within -1.0 to +1.0	Excellent.
Between ± 1.0 to ± 2.0	Good.
Between ± 2.0 to ± 3.0	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

Page 1 of 1

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

Contact details:

Email: clinqc@cmcvellore.ac.in

Contact Number: 0416-2283102

Pamela Christudoss

Dr. Pamela Christudoss
 CMC EQAS Co-Ordinator
 Christian Medical College, Vellore

***** End of Report *****