

LAB MONTHLY SUMMARY

Lab Name **JEEVAN DIAGNOSTIC CENTRE**Lab No **13184**Month **May**Year **2020**Constituent Group **Chemistry II**Date of Result Entered : **01/07/2020**Date of Report Published : **04/08/2020**

| Sl.No | Analyte | Method / Principle Name | Analyzer Name | No of Participants | DV | Participants | | Your Value | SDI | U |
|-------|----------------|--|---------------|--------------------|--------|--------------|-------|-------------|-------|------|
| | | | | | | CV | SD | | | |
| 1 | GLUCOSE I | GOD-POD | Any Analyser | 858 | 105.04 | 10.46 | 10.98 | 101 mg/dl | -0.37 | 0.75 |
| 2 | UREA I | UREASE UV / GLDH | Any Analyser | 517 | 29.24 | 16.17 | 4.73 | 39.8 mg/dl | 2.23 | 0.42 |
| 3 | CREATININE I | OTHERS (Any Other Principles / Methods) | Any Analyser | 93 | 4.60 | 12.62 | 0.58 | 5.1 mg/dl | 0.86 | 0.12 |
| 4 | T-PROTEIN I | BIURET - colorimetric | Any Analyser | 639 | 5.23 | 9.21 | 0.48 | 6 g/dl | 1.60 | 0.04 |
| 5 | ALBUMIN I | BCG - colorimetric | Any Analyser | 635 | 3.20 | 10.09 | 0.32 | 3.6 g/dl | 1.24 | 0.03 |
| 6 | URIC ACID I | ENZYMATIC / URICASE Colorimetric | Any Analyser | 609 | 6.34 | 14.77 | 0.94 | 6.8 mg/dl | 0.49 | 0.08 |
| 7 | CHOLESTEROL I | CHOD-PAP | Any Analyser | 724 | 108.65 | 11.96 | 12.99 | 118.6 mg/dl | 0.77 | 0.97 |
| 8 | TRIGLYCERIDE I | GPO-PAP / Enzymatic Colorimetric / End Point | Any Analyser | 652 | 235.57 | 12.43 | 29.29 | 271 mg/dl | 1.21 | 2.29 |

| SDI Range | Interpretation |
|--------------------------------|--|
| Within -1.0 to +1.0 | Excellent. |
| Between ± 1.0 to ± 2.0 | Good. |
| Between ± 2.0 to ± 3.0 | Accept with caution. Warning Signal. |
| Beyond ± 3.0 | Unacceptable performance. Action Signal. |

Page 1 of 1

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

Contact details:

Email: clinqc@cmcvellore.ac.in

Contact Number: 0416-2283102

Pamela Christudoss

Dr. Pamela Christudoss
 CMC EQAS Co-Ordinator
 Christian Medical College, Vellore

***** End of Report *****