



Ace Medical Solutions (India) Pvt. Ltd.

No : 3/751, 1st Cross Street, Surveyor Colony, K.Pudur, Madurai - 625 007.

Phone : 0452 - 2561691 Mobile : +91 94434 74610



GSTIN: 33AAHCA8053H1ZR • DL No. MDU/4205/20B,MDU/3971/21B • CIN : U33110TN2009PTC072140

CALIBRATION CERTIFICATE

Customer Name and Address: M/s. National lab,

Narimedu, Madurai.

Date :06/07/2019

Certificate No	Calibrated On	Calibration Due On	Page No
2886162	05/07/2019	04/07/2020	1 of 2

Details of Device under Calibration(DUC):

DUC : Semi Auto Analyzer	Type: NA
Make: Erba	Range: NA
Model: Chem - 5 Plus	Least Count:NA
Serial No:NA	DUC Condition on Receipt:OK
ID.No: NL/LAB/SAA/001	Department: Laboratory

Room Temperature:25°C±2 Room Humidity:30-60% (At the time of Calibration)

Details of Test Equipment Used

Test Equipment	Manufacturer	Model	Serial No
Electrical Safety Analyzer	Rigel	Safe test 60	41H-0958
Thermo Hygrometer	HTC	1	----


Calibrated At	Traceability	Certificate No	Cal.Date
HS Calibration, Coimbatore	PSG College of Technology	3879	21/01/2019
Quality Concepts, Madurai	Scientific & Industrial Testing & Research Centre	20190218	11/02/2019

ELECTRICAL SAFETY: Please check the attached log sheet


Note:

1. Calibration Points were selected as per customer request.
2. Each reading were taken by Repeatability Test(i.e.)each reading in a row is a mean of 10 readings.
3. Results Reported are valid at the time of and under the stated conditions of measurements.

Calibrated By:


(Calibration Engineer)

Approved By:


(Technical Manager)

