



Ace Medical Solutions (India) Pvt. Ltd.

No : 3/751, 1st Cross Street, Surveyor Colony, K.Pudur, Madurai - 625 007.

Phone : 0452 - 2561691 Mobile : +91 94434 74610



GSTIN. 33AAHCA8053H1ZR • DL No. MDU/4205/20B,MDU/3971/21B • CIN : U33110TN2009PTC072140

CALIBRATION CERTIFICATE

Customer Name & Address: M/s. National Lab,
Narimedu, Madurai.

Date: 06/07/2019

Certificate No	Calibrated On	Calibration Due On	Page No
2886156	05/07/2019	04/07/2020	1 of 1

Details of Device under Calibration (DUC):

DUC: Centrifuge	Type: Electric
Make: Remi	Range: 0-4000rpm
Model: R-8C	Least Count: NA
SI. No: HGCL-10552	DUC Condition on Receipt: Ok
Id. No: VL/LAB/CEN/001	Department: Laboratory

Room Temperature: 25°C±2 Room Humidity: 30-60% (At the time of Calibration)

Details of Test Equipment Used

Test Equipment	Manufacturer	Model	Serial No
Tachometer	LUTRON	DT-2234C	L335794
Thermo Hygrometer	HTC	1	----

Calibrated At	Traceability	Certificate No	Date
Faber Sindoori Management services, Chennai	National Institute of Standards and Technology	FSMS/CL/05/18-08/071-01	06/08/2018
Quality Concepts, Madurai	Scientific & Industrial Testing & Research Centre	20190218	11/02/2019

Results:

SI. No	DUC Reading(rpm)	Standard Reading(rpm)	Correction(rpm)
1	1500	1490	10
2	2500	2510	10
3	3000	3100	100

Note:

1. Calibration Points were selected as per customer request.
2. Each reading were taken by Repeatability Test (i.e.) each reading in a row is a mean of 10 readings.
3. Results Reported are valid at the time of and under the stated conditions of measurements.

Calibrated By:

(Calibration Engineer)

Approved by:

(Technical Manager)

