



Ace Medical Solutions (India) Pvt. Ltd.

No : 3/751, 1st Cross Street, Surveyor Colony, K.Pudur, Madurai - 625 007.

Phone : 0452 - 2561691 Mobile : +91 94434 74610



GSTIN: 33AAHCA8053H12R • DL No. MDU/4205/208, MDU/3971/218 • CIN : U33110TN2009PTC072140

CALIBRATION CERTIFICATE

Customer Name & Address: M/s. National Lab,
Narimedu, Madurai.

Date: 06/07/2019

Certificate No	Calibrated On	Calibration Due On	Page No
2886159	05/07/2019	04/07/2020	1 of 1

Details of Device under Calibration (DUC):

DUC: Micropipette	Type: Variable
Make: NA	Range: 0-50µl
Model: NA	Least Count: NA
Sl. No: YE3K041320	DUC Condition on Receipt: OK
Id. No: NL/LAB/MP/003	Department: Laboratory

Room Temperature: 25°C±2 Room Humidity: 30-60% (At the time of Calibration)

Details of Test Equipment Used

Test Equipment	Manufacturer	Model	Serial No
Electronic weighing scale	Pentagon	EPS10T	E10T151189
Thermo Hygrometer	HTC	I	---

Calibrated At	Traceability	Certificate No	Date
Pentagon Electronic Scale Service	Government of Tamilnadu	1181631	16/03/2018
Quality Concepts	Scientific & Industrial Testing & Research Centre	20190218	11/02/2019

Results:

Sl. No	DUC Reading(µl)	Standard Reading(µl)		Correction (µl)
		Trial 1	Trial 2	
1	50	51	50	1

Note:

1. Allowable Tolerance: ±1 µl
2. Calibration Points were selected as per customer request.
3. Each reading were taken by Repeatability Test (i.e.) each reading in a row is a mean of 10 readings.
4. Results Reported are valid at the time of and under the stated conditions of measurements.

Calibrated By:

[Signature]

(Calibration Engineer)

Approved by:

[Signature]

(Technical Manager)

