



**REAGENT
STORAGE**



Ace Medical Solutions (India) Pvt. Ltd.

No : 3/751, 1st Cross Street, Surveyor Colony, K.Pudur, Madurai - 625 007.
Phone : 0452 - 2561691 Mobile : +91 94434 74610



GSTIN: 33AAHCA8053H1ZR • DL No. MDU/4205/208,MDU/3971/21B • CIN : U33110TN2009PTC072140

CALIBRATION CERTIFICATE

Customer Name & Address: M/s. National Lab,
Narimedu, Madurai.

Date: 06/07/2019

Certificate No	Calibrated On	Calibration Due On	Page No
2886165	05/07/2019	04/07/2020	1 of 1

Details of Device under Calibration (DUC):

DUC: Refrigerator	Type: Direct Cool
Make: LG	Range: NA
Model: GL-245BM5/2012	Least Count: NA
SI. No: NA	DUC Condition on Receipt: Ok
Id. No: SBB-LAB-II-RF-001	Department: Laboratory

Environmental Conditions: Temperature: 20°C±2°C Humidity: 30-60%

Details of Test Equipment Used

Test Equipment	Manufacturer	Model	Serial No
RTD Thermometer	CENTER	375	140811115
Thermo Hygrometer	HTC	1	---

Calibrated At	Traceability	Certificate No	Cal. Date
Quality Concepts, Madurai	Scientific & Industrial Testing & Research centre	20181121	10/02/2018
Quality Concepts, Madurai	Scientific & Industrial Testing & Research Centre	20181120	12/02/2018


Results:

SI. No	DUC Reading(°C)	Standard Reading(°C)	Correction (°C)
1	5.2	5.2	0


Note:

1. Calibration Points were selected as per customer request.
2. Each reading were taken by Repeatability Test (i.e.) each reading in a row is a mean of 10 readings.
3. Results Reported are valid at the time of and under the stated conditions of measurements.

Calibrated By:


(Calibration Engineer)

Approved by:


(Technical Manager)

