



PROFICIENCY TESTING REPORT
ISHTM-AIIMS EXTERNAL QUALITY ASSURANCE PROGRAMME
 NABL accredited program as per ISO/IEC 17043:2010 standard
 Organized By Department of Hematology, AIIMS, New Delhi-110029



Duration of stability testing - minimum upto 8 days at ambient temp. after dispatch of specimens

EQAP CODE No. : 2930

Distribution No.: 149-G

Month/Year: December/2019

Instrument ID: SYSMEX SN 11082

Name & Contact No. of PT Co-ordinator: Dr. Seema Tyagi (Prof.), Hematology, AIIMS, Delhi,
 Tel: 9013085730 , E-Mail : accuracy2000@gmail.com

Date of issue & status of the report: 09-03-2020[Final].

CBC and Retic Assessment

Test Parameters	S.No.	Among Lab (Accuracy Testing)						Within Lab (Precision Testing)			
		Your Result 1	Your Result 2	Your Results Sum of 2 Value	Consensus result sum of 2 values (Assigned Value)	Uncertainty of Assigned Values	Z Score	Yours Results Diff. of 2 Values	Consensus Result Diff. of 2 values (Assigned Value)	Uncertainty of Assigned Values	Z Score
WBC x10 ³ /µl	1	5.4	5.1	10.5	10.1	0.0310	0.53	0.3	0.1	0.0740	1.80
RBC x10 ⁶ /µl	1	4.42	4.4	8.82	8.64	0.0090	0.72	0.02	0.04	0.0020	-0.45
Hb g/dl	1	12.6	12.5	25.1	26.1	0.0290	-1.50	0.1	0.1	0.0080	0.00
HCT%	1	38	37.9	75.9	76.75	0.1550	-0.21	0.1	0.4	0.0250	-0.81
MCV-fl	1	85.8	85.7	171.5	177.7	0.2910	-0.77	0.1	0.3	0.0250	-0.45
MCH-Pg	1	28.5	28.4	56.9	60.5	0.0820	-1.89	0.1	0.3	0.0220	-0.67
MCHC-g/dl	1	33.2	33.1	66.3	68.1	0.1450	-0.46	0.1	0.3	0.0200	-0.67
Plt. x10 ³ /µl	1	170	159	329	338	1.05	-0.35	11	5	0.31	1.35
Retic %	2	3.6	3.4	7	7	0.19	0.00	0.2	0.3	0.01	-0.45

P.S . Assesment

YOUR REPORT			CONSENSUS REPORT		
DLC%	3	Nrbcs=6 - 8 , Poly=34 / 12 L=03 / 00, E=03, Mono/Promono=04 / 00 , B1=02 P.M.=00, Mye=34, Meta=06, Other=BASOPHIL 02	Poly: 50-60, Lymph: 2-6, nRBC/Eo/Mono/Blast/Pro: 0-5, My: 10-20, Meta: 10-15		
RBC Morphology	3	NCNC TO MCHC ANISO ++ POI..++	Predominantly: Normocytic Normochromic. Moderate: Micro. Mild: Hypo, Aniso.		
Diagnosis	3	CML - CP CHRONIC MYELOID LEUKAEMIA. CHRONIC PHASE	Chronic Myeloid Leukemia (Chronic Phase) : CML-CP		

EQAP Code No.:
2930

Distribution No.: 149-G Month/Year: December/2019

Instrument ID: SYSMEX SN
11082**COMBINED DATA VALUES OF TOTAL PARTICIPANTS**

Test parameters	S.No.	Total participants covered in the current dist.	Total No. responded	% of Labs with Z Score 0-2		% of Labs with Z Score 2-3		% of Labs with Z Score >3	
				Among labs	Within lab	Among labs	Within lab	Among labs	Within lab
WBC x10 ³ /µl	1	450	278	87.77	88.13	5.04	4.68	7.19	7.19
RBC x10 ⁶ /µl	1	450	278	89.21	93.88	6.12	1.8	4.68	3.6
Hb g/dl	1	450	278	84.89	88.85	8.63	6.12	6.47	5.04
HCT%	1	450	278	91.01	92.81	7.55	4.32	1.44	2.88
MCV-fl	1	450	278	94.24	93.17	3.96	3.6	1.44	2.88
MCH-Pg	1	450	278	85.97	90.29	7.91	5.76	6.12	3.6
MCHC-g/dl	1	450	278	93.53	88.85	5.4	5.76	1.08	5.4
Plt. x10 ³ /µl	1	450	278	88.49	91.73	6.83	5.76	4.68	2.52
ReticCount%	2	450	230	93.91	82.17	2.61	13.48	3.48	6.52
PS Assessment	3	450	254	Acceptable:95.8%,Warning Signal:3.4%,Unacceptable :0.8%					

Comments:*1). Among Lab (EQA) : Results acceptable.****2). Within Lab (IQA) : Precision acceptable.****Note-1: EQA** (External Quality Assurance) : Your Performance among various of participating labs in PT, to determine the accuracy of your results.**IQA** (Internal Quality Assurance) : Your Performance of comparison of two consecutive measurement values within your lab to test the precision of your autoanalyzer.**Note-2:** Z score among & within lab were calculated, as per to ISO/IEC 13528:2015 standard. Z score among lab (EQA)= (Your Result Sum of two values - Consensus Result sum of two values)/(Normalised IQR)

Z score within lab (IQA)= (Your Result Difference of two values - Consensus Result difference of two values)/(Normalised IQR)

IQR = Quartile 3 - Quartile 1 of participant data, Normalised IQR = 0.7413 x IQR

Note-3: Z score 0 to ±2: Acceptable, Z score ±2 to ±3 :Warning Signal, Z score > ±3 : Unacceptable [As per ISO/IEC 13528:2015 standard]**Note-4:** Z score value between "0 to ±2" are texted in green colour. Z score value between "±2 to ±3" are texted in orange colour. Z score value > ±3 are texted in red colour.**Note-5:** Homogeneity and stability testing of PT sample were done as per ISO 13528:2015 standard. To pass homogeneity test, between sample SD (Ss) should be smaller than the check value (0.3*SDPA). To pass the stability test, average difference in measurement values of first and last day sample ($\bar{x}-\bar{y}$) should be smaller than the check value (0.3*SDPA).**Note-6:** ISHTM-AIIMS-EQAP does not subcontract any task of its scheme**Note-7:** Participants are free to use methods/analyzer of their own choice.**Note-8:** Proficiency testing (PT) samples are sent quarterly to each participant.**Note-9:** All the necessary details regarding design and implementation of PT, are provided in the instruction sheet as well as on programme's website www.ishtmaiimseqap.com.

Report authorized by,



Dr. R. Saxena

Prof & Head, Hematology, AIIMS, Delhi.

PT Co-ordinator: ISHTM-AIIMS-EQAP

-----End Of Report-----