

LAB MONTHLY SUMMARY

Lab Name AB DIAGNOSTICS

Lab No 11537

Month June

Year 2020

Constituent Group Chemistry II

Date of Result Entered : 19/06/2020

Date of Report Published : 08/08/2020

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE I	GOD-POD	Mindray	113	80.48	9.61	7.73	72 mg/dl	-1.10	1.46
2	UREA I	NED DYE	Any Analyser	85	49.69	10.76	5.35	49.6 mg/dl	-0.02	1.16
3	CREATININE I	JAFFE RATE-BLANKED & COMPENSATED/Alkaline Picrate	Mindray	80	3.44	12.16	0.42	3.1 mg/dl	-0.81	0.09
4	T.BILIRUBIN I	DIAZONIUM SALT (Colorimetric) /JENDRASSIK	Mindray	77	5.35	15.08	0.81	5.8 mg/dl	0.56	0.18
5	T-PROTEIN I	BIURET - colorimetric	Mindray	87	4.82	8.84	0.43	5.2 g/dl	0.89	0.09
6	ALBUMIN I	BCG - colorimetric	Mindray	79	2.97	9.13	0.27	3.2 g/dl	0.85	0.06
7	URIC ACID I	ENZYMATIC / URICASE Colorimetric	Mindray	86	4.30	18.87	0.81	3.9 mg/dl	-0.49	0.17
8	CHOLESTEROL II	CHOD-PAP	Mindray	104	96.86	12.30	11.92	79.5 mg/dl	-1.46	2.34
9	TRIGLYCERIDE I	GPO-PAP / Enzymatic Colorimetric / End Point	Mindray	98	88.22	16.90	14.91	172.8 mg/dl	5.67	3.01

SDI Range	Interpretation
Within -1.0 to +1.0	Excellent.
Between ±1.0 to ±2.0	Good.
Between ±2.0 to ±3.0	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

Contact details:

Email: [clinqc@cmcvellore.ac.in](mailto:clinqc@cmcvellore.ac.in)

Contact Number: 0416-2283102

*Pamela Christudoss*

Dr. Pamela Christudoss

CMC EQAS Co-Ordinator

Christian Medical College, Vellore

\*\*\*\*\* End of Report \*\*\*\*\*