

EW LAB MONTHLY SUMMARY

Lab Name	JOTHI X-RAY AND LABORATORY	Details About Robust Analysis
Lab No	9661	Detail About Monthly Summary
Month	September	Detail about SDI
Year	2020	
Constituent Group	HbA1c	

[Click on the analyte to view Graphical Data](#)

Date of Result Entered : 25/09/2020

Date of Report Published : 07/10/2020

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	HbA1c	HPLC-ION EXCHANGE	Biorad (D10 / D100 / Variant Turbo)	473	5.73	7.35	0.42	5.9 %	0.40	0.04

SDI Range	Interpretation
Within -1.0 to +1.0	Excellent.
Between ±1.0 to ±2.0	Good.
Between ±2.0 to ±3.0	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links

Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out