

## LAB MONTHLY SUMMARY

Lab Name **RAGHAV PATH LABS PVT LTD** Lab No **6902**  
 Month **August** Year **2020**  
 Constituent **Chemistry II**  
 Group

Date of Result Entered : 10/08/2020

Date of Report Published : 10/09/2020

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE I	HEXOKINASE	Beckman AU480/680/5800/DXC700AU	17	115.44	3.03	3.50	90.5 mg/dl	-7.12	1.70
2	UREA I	UREASE UV / GLDH	Beckman AU480/680/5800/DXC700AU	26	74.85	5.54	4.15	78.6 mg/dl	0.90	1.63
3	CREATININE I	JAFFE RATE-BLANKED &COMPENSATED/Alkaline Picrate	Beckman AU480/680/5800/DXC700AU	21	9.52	8.74	0.83	9.6 mg/dl	0.10	0.36
4	T.BILIRUBIN I	DIAZONIUM SALT (Colorimetric) /JENDRASSIK	Beckman AU480/680/5800/DXC700AU	27	6.21	4.40	0.27	6.7 mg/dl	1.79	0.11
5	T-PROTEIN I	BIURET - colorimetric	Beckman AU480/680/5800/DXC700AU	25	4.70	3.23	0.15	4.3 g/dl	-2.63	0.06
6	ALBUMIN I	BCG - colorimetric	Beckman AU480/680/5800/DXC700AU	26	2.90	3.41	0.10	2.8 g/dl	-1.01	0.04
7	URIC ACID I	ENZYMATIC / URICASE Colorimetric	Beckman AU480/680/5800/DXC700AU	21	5.61	5.54	0.31	6.6 mg/dl	3.18	0.14
8	CHOLESTEROL I	CHOD-PAP	Beckman AU480/680/5800/DXC700AU	26	101.75	3.38	3.44	101.7 mg/dl	-0.01	1.35
9	TRIGLYCERIDE I	GPO-PAP / Enzymatic Colorimetric / End Point	Beckman AU480/680/5800/DXC700AU	23	285.62	5.25	14.99	262.3 mg/dl	-1.56	6.25

SDI Range	Interpretation
Within -1.0 to +1.0	Excellent.
Between $\pm 1.0$ to $\pm 2.0$	Good.
Between $\pm 2.0$ to $\pm 3.0$	Accept with caution. Warning Signal.
Beyond $\pm 3.0$	Unacceptable performance. Action Signal.

Page 1 of 1

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

Contact details:

Email: clinqc@cmcvellore.ac.in

Contact Number: 0416-2283102



Dr. Pamela Christudoss  
 CMC EQAS Co-Ordinator  
 Christian Medical College, Vellore

\*\*\*\*\* End of Report \*\*\*\*\*