

LAB MONTHLY SUMMARY

Lab Name **RAGHAV PATH LABS PVT LTD** Lab No **6902**
 Month **July** Year **2020**
 Constituent **Chemistry II**
 Group

Date of Result Entered : 16/07/2020

Date of Report Published : 14/08/2020

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE I	HEXOKINASE	Beckman AU480/680/5800/DXC700AU	18	418.23	5.75	24.04	343.2 mg/dl	-3.12	11.33
2	UREA I	UREASE UV / GLDH	Beckman AU480/680/5800/DXC700AU	28	95.45	3.51	3.35	95.6 mg/dl	0.04	1.27
3	CREATININE I	JAFFE RATE-BLANKED &COMPENSATED/Alkaline Picrate	Beckman AU480/680/5800/DXC700AU	22	0.89	11.80	0.10	0.9 mg/dl	0.10	0.04
4	T.BILIRUBIN I	DIAZONIUM SALT (Colorimetric) /JENDRASSIK	Beckman AU480/680/5800/DXC700AU	27	4.30	3.91	0.17	4.2 mg/dl	-0.60	0.06
5	T-PROTEIN I	BIURET - colorimetric	Beckman AU480/680/5800/DXC700AU	27	4.82	3.49	0.17	5 g/dl	1.07	0.06
6	ALBUMIN I	BCG - colorimetric	Beckman AU480/680/5800/DXC700AU	28	2.99	4.65	0.14	3 g/dl	0.07	0.05
7	URIC ACID I	ENZYMATIC / URICASE Colorimetric	Beckman AU480/680/5800/DXC700AU	24	7.33	5.33	0.39	7.6 mg/dl	0.69	0.16
8	CHOLESTEROL I	CHOD-PAP	Beckman AU480/680/5800/DXC700AU	27	101.58	4.81	4.89	105 mg/dl	0.70	1.88
9	TRIGLYCERIDE I	GPO-PAP / Enzymatic Colorimetric / End Point	Beckman AU480/680/5800/DXC700AU	25	226.62	3.36	7.62	224.7 mg/dl	-0.25	3.05

SDI Range	Interpretation
Within -1.0 to +1.0	Excellent.
Between ±1.0 to ±2.0	Good.
Between ±2.0 to ±3.0	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Page 1 of 1

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

Contact details:

Email:clinqc@cmcvellore.ac.in

Contact Number: 0416-2283102

Pamela Christudoss

Dr. Pamela Christudoss
 CMC EQAS Co-Ordinator
 Christian Medical College, Vellore

***** End of Report *****