

LAB MONTHLY SUMMARY

Lab Name **RAGHAV PATH LABS PVT LTD** Lab No **6902**
 Month **June** Year **2020**
 Constituent **Chemistry II**
 Group

Date of Result Entered : 24/06/2020

Date of Report Published : 08/08/2020

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE I	HEXOKINASE	Beckman AU480/680/5800/DXC700AU	17	77.44	5.48	4.24	61.4 mg/dl	-3.78	2.06
2	UREA I	UREASE UV / GLDH	Beckman AU480/680/5800/DXC700AU	27	50.50	4.27	2.16	50.6 mg/dl	0.05	0.83
3	CREATININE I	JAFFE RATE-BLANKED &COMPENSATED/Alkaline Picrate	Beckman AU480/680/5800/DXC700AU	21	3.67	6.10	0.22	3.9 mg/dl	1.03	0.10
4	T.BILIRUBIN I	DIAZONIUM SALT (Colorimetric) /JENDRASSIK	Beckman AU480/680/5800/DXC700AU	27	5.75	5.58	0.32	5.4 mg/dl	-1.09	0.12
5	T-PROTEIN I	BIURET - colorimetric	Beckman AU480/680/5800/DXC700AU	26	4.28	3.16	0.14	4.3 g/dl	0.15	0.05
6	ALBUMIN I	BCG - colorimetric	Beckman AU480/680/5800/DXC700AU	27	2.64	3.75	0.10	2.7 g/dl	0.61	0.04
7	URIC ACID I	ENZYMATIC / URICASE Colorimetric	Beckman AU480/680/5800/DXC700AU	23	3.88	4.65	0.18	3.7 mg/dl	-1.00	0.08
8	CHOLESTEROL I	CHOD-PAP	Beckman AU480/680/5800/DXC700AU	27	87.68	7.50	6.58	93 mg/dl	0.81	2.53
9	TRIGLYCERIDE I	GPO-PAP / Enzymatic Colorimetric / End Point	Beckman AU480/680/5800/DXC700AU	25	81.88	6.98	5.71	74.9 mg/dl	-1.22	2.28

SDI Range	Interpretation
Within -1.0 to +1.0	Excellent.
Between ± 1.0 to ± 2.0	Good.
Between ± 2.0 to ± 3.0	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

Page 1 of 1

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

Contact details:

Email:clinqc@cmcvellore.ac.in

Contact Number: 0416-2283102



Dr. Pamela Christudoss
 CMC EQAS Co-Ordinator
 Christian Medical College, Vellore

***** End of Report *****