

LAB MONTHLY SUMMARY

Lab Name **RAGHAV PATH LABS PVT LTD** Lab No **6902**
 Month **May** Year **2020**
 Constituent **Chemistry II**
 Group

Date of Result Entered : 24/06/2020

Date of Report Published : 04/08/2020

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE I	HEXOKINASE	Beckman AU480/680/5800/DXC700AU	15	106.40	4.55	4.84	80.4 mg/dl	-5.37	2.50
2	UREA I	UREASE UV / GLDH	Beckman AU480/680/5800/DXC700AU	25	26.68	4.57	1.22	27.8 mg/dl	0.92	0.49
3	CREATININE I	JAFFE RATE-BLANKED &COMPENSATED/Alkaline Picrate	Beckman AU480/680/5800/DXC700AU	23	4.91	6.70	0.33	5.6 mg/dl	2.10	0.14
4	T.BILIRUBIN I	DIAZONIUM SALT (Colorimetric) /JENDRASSIK	Beckman AU480/680/5800/DXC700AU	23	3.00	3.93	0.12	3.1 mg/dl	0.85	0.05
5	T-PROTEIN I	BIURET - colorimetric	Beckman AU480/680/5800/DXC700AU	25	4.86	1.15	0.06	4.9 g/dl	0.71	0.02
6	ALBUMIN I	BCG - colorimetric	Beckman AU480/680/5800/DXC700AU	24	2.97	3.94	0.12	3.2 g/dl	1.97	0.05
7	URIC ACID I	ENZYMATIC / URICASE Colorimetric	Beckman AU480/680/5800/DXC700AU	21	6.05	5.30	0.32	5.8 mg/dl	-0.78	0.14
8	CHOLESTEROL I	CHOD-PAP	Beckman AU480/680/5800/DXC700AU	24	106.58	6.65	7.08	116 mg/dl	1.33	2.89
9	TRIGLYCERIDE I	GPO-PAP / Enzymatic Colorimetric / End Point	Beckman AU480/680/5800/DXC700AU	20	252.69	3.29	8.33	247.4 mg/dl	-0.64	3.72

SDI Range	Interpretation
Within -1.0 to +1.0	Excellent.
Between ± 1.0 to ± 2.0	Good.
Between ± 2.0 to ± 3.0	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

Page 1 of 1

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

Contact details:

Email: clinqc@cmcvellore.ac.in

Contact Number: 0416-2283102



Dr. Pamela Christudoss
 CMC EQAS Co-Ordinator
 Christian Medical College, Vellore

***** End of Report *****