

LAB MONTHLY SUMMARY

Lab Name **RAGHAV PATH LABS PVT LTD** Lab No **6902**
 Month **September** Year **2020**
 Constituent **Chemistry II**
 Group

Date of Result Entered : 03/09/2020

Date of Report Published : 07/10/2020

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE I	HEXOKINASE	Beckman AU480/680/5800/DXC700AU	19	233.82	4.17	9.75	226.4 mg/dl	-0.76	4.47
2	UREA I	UREASE UV / GLDH	Beckman AU480/680/5800/DXC700AU	26	134.70	6.27	8.45	163.4 mg/dl	3.4	3.31
3	CREATININE I	JAFFE RATE-BLANKED &COMPENSATED/Alkaline Picrate	Beckman AU480/680/5800/DXC700AU	24	1.44	7.07	0.10	1.6 mg/dl	1.57	0.04
4	T.BILIRUBIN I	DIAZONIUM SALT (Colorimetric) /JENDRASSIK	Beckman AU480/680/5800/DXC700AU	29	5.38	5.63	0.30	6 mg/dl	2.05	0.11
5	T-PROTEIN I	BIURET - colorimetric	Beckman AU480/680/5800/DXC700AU	27	4.76	5.15	0.24	5 g/dl	0.98	0.09
6	ALBUMIN I	BCG - colorimetric	Beckman AU480/680/5800/DXC700AU	28	3.03	4.16	0.13	3.1 g/dl	0.56	0.05
7	URIC ACID I	ENZYMATIC / URICASE Colorimetric	Beckman AU480/680/5800/DXC700AU	25	7.13	9.99	0.71	6.8 mg/dl	-0.46	0.28
8	CHOLESTEROL I	CHOD-PAP	Beckman AU480/680/5800/DXC700AU	26	101.67	4.95	5.03	102.6 mg/dl	0.18	1.97
9	TRIGLYCERIDE I	GPO-PAP / Enzymatic Colorimetric / End Point	Beckman AU480/680/5800/DXC700AU	26	88.77	6.81	6.04	76.6 mg/dl	-2.01	2.37

SDI Range	Interpretation
Within -1.0 to +1.0	Excellent.
Between ± 1.0 to ± 2.0	Good.
Between ± 2.0 to ± 3.0	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

Page 1 of 1

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

Contact details:

Email:clinqc@cmcvellore.ac.in

Contact Number: 0416-2283102



Dr. Pamela Christudoss
 CMC EQAS Co-Ordinator
 Christian Medical College, Vellore

***** End of Report *****