

LAB MONTHLY SUMMARY

Lab Name **SRI KUMARAN HOSPITAL**Lab No **11392**Month **September**Year **2020**Constituent Group **Chemistry II**Date of Result Entered : **25/09/2020**Date of Report Published : **07/10/2020**

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE I	GOD-POD	Biosystems	209	228.65	9.52	21.76	237 mg/dl	0.38	3.01
2	UREA I	UREASE BERTHELOT	Biosystems	57	130.20	15.14	19.71	130 mg/dl	-0.01	5.22
3	CREATININE I	JAFFE RATE-BLANKED & COMPENSATED/Alkaline Picrate	Biosystems	192	1.40	11.82	0.17	1.4 mg/dl	0.00	0.02
4	T.BILIRUBIN I	DIAZONIUM SALT (Colorimetric) /JENDRASSIK	Biosystems	157	5.42	12.87	0.70	5.3 mg/dl	-0.17	0.11
5	T-PROTEIN I	BIURET - colorimetric	Biosystems	178	5.27	8.69	0.46	5.2 g/dl	-0.15	0.07
6	ALBUMIN I	BCG - colorimetric	Biosystems	180	3.26	8.75	0.28	3.6 g/dl	1.19	0.04
7	URIC ACID I	ENZYMATIC / URICASE Colorimetric	Biosystems	172	7.35	12.23	0.90	8 mg/dl	0.72	0.14
8	CHOLESTEROL I	CHOD-PAP	Biosystems	186	107.65	11.15	12.00	107 mg/dl	-0.05	1.76
9	TRIGLYCERIDE I	GPO-PAP / Enzymatic Colorimetric / End Point	Biosystems	166	100.44	13.48	13.54	92 mg/dl	-0.62	2.10

SDI Range	Interpretation
Within -1.0 to +1.0	Excellent.
Between ± 1.0 to ± 2.0	Good.
Between ± 2.0 to ± 3.0	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

Page 1 of 1

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

Contact details:

Email: clinqc@cmcvellore.ac.in

Contact Number: 0416-2283102

Pamela Christudoss

Dr. Pamela Christudoss

CMC EQAS Co-Ordinator

Christian Medical College, Vellore

***** End of Report *****