

LAB MONTHLY SUMMARY

Lab Name **SRI KUMARAN HOSPITAL** Lab No **11392**
 Month **August** Year **2020**
 Constituent Group **Chemistry II**

Date of Result Entered : 25/08/2020

Date of Report Published : 10/09/2020

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE I	GOD-POD	Biosystems	221	112.95	9.75	11.01	111 mg/dl	-0.18	1.48
2	UREA I	UREASE BERTHELOT	Biosystems	58	72.89	13.75	10.02	79 mg/dl	0.61	2.63
3	CREATININE I	JAFFE RATE-BLANKED & COMPENSATED/Alkaline Picrate	Biosystems	197	8.96	10.43	0.93	8.9 mg/dl	-0.06	0.13
4	T.BILIRUBIN I	DIAZONIUM SALT (Colorimetric) /JENDRASSIK	Biosystems	150	6.20	11.29	0.70	7 mg/dl	1.14	0.11
5	T-PROTEIN I	BIURET - colorimetric	Biosystems	171	5.35	9.02	0.48	5.7 g/dl	0.73	0.07
6	ALBUMIN I	BCG - colorimetric	Biosystems	169	3.21	8.79	0.28	3.5 g/dl	1.03	0.04
7	URIC ACID I	ENZYMATIC / URICASE Colorimetric	Biosystems	145	6.20	11.30	0.70	6 mg/dl	-0.29	0.12
8	CHOLESTEROL I	CHOD-PAP	Biosystems	184	107.04	9.28	9.93	108 mg/dl	0.10	1.46
9	TRIGLYCERIDE I	GPO-PAP / Enzymatic Colorimetric / End Point	Biosystems	150	276.86	9.68	26.81	290 mg/dl	0.49	4.38

SDI Range	Interpretation
Within -1.0 to +1.0	Excellent.
Between ± 1.0 to ± 2.0	Good.
Between ± 2.0 to ± 3.0	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

Page 1 of 1

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

Contact details:

Email: clinqc@cmcvellore.ac.in

Contact Number: 0416-2283102



Dr. Pamela Christudoss
 CMC EQAS Co-Ordinator
 Christian Medical College, Vellore

***** End of Report *****