

## LAB MONTHLY SUMMARY

Lab Name **SRI KUMARAN HOSPITAL** Lab No **11392**  
 Month **July** Year **2020**  
 Constituent Group **Chemistry II**

Date of Result Entered : 23/07/2020

Date of Report Published : 14/08/2020

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE I	GOD-POD	Biosystems	208	391.06	8.03	31.39	396 mg/dl	0.16	4.35
2	UREA I	UREASE BERTHELOT	Biosystems	62	89.64	12.86	11.53	103 mg/dl	1.16	2.93
3	CREATININE I	JAFFE RATE-BLANKED & COMPENSATED/Alkaline Picrate	Biosystems	189	0.94	11.85	0.11	0.9 mg/dl	-0.36	0.02
4	T.BILIRUBIN I	DIAZONIUM SALT ( Colorimetric ) /JENDRASSIK	Biosystems	161	4.28	13.33	0.57	4.3 mg/dl	0.04	0.09
5	T-PROTEIN I	BIURET - colorimetric	Biosystems	165	5.32	9.17	0.49	5.3 g/dl	-0.04	0.08
6	ALBUMIN I	BCG - colorimetric	Biosystems	169	3.25	9.74	0.32	3.5 g/dl	0.79	0.05
7	URIC ACID I	ENZYMATIC / URICASE Colorimetric	Biosystems	165	7.67	11.54	0.88	8.2 mg/dl	0.60	0.14
8	CHOLESTEROL I	CHOD-PAP	Biosystems	190	107.26	10.02	10.75	98 mg/dl	-0.86	1.56
9	TRIGLYCERIDE I	GPO-PAP / Enzymatic Colorimetric / End Point	Biosystems	175	219.96	13.07	28.75	239 mg/dl	0.66	4.35

SDI Range	Interpretation
Within -1.0 to +1.0	Excellent.
Between $\pm 1.0$ to $\pm 2.0$	Good.
Between $\pm 2.0$ to $\pm 3.0$	Accept with caution. Warning Signal.
Beyond $\pm 3.0$	Unacceptable performance. Action Signal.

Page 1 of 1

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

Contact details:

Email:clinqc@cmcvellore.ac.in

Contact Number: 0416-2283102



Dr. Pamela Christudoss  
 CMC EQAS Co-Ordinator  
 Christian Medical College, Vellore

\*\*\*\*\* End of Report \*\*\*\*\*