

LAB MONTHLY SUMMARY

Lab Name **SRI KUMARAN HOSPITAL** Lab No **11392**
 Month **June** Year **2020**
 Constituent Group **Chemistry II**

Date of Result Entered : 02/07/2020

Date of Report Published : 08/08/2020

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE I	GOD-POD	Biosystems	208	80.18	10.16	8.15	75 mg/dl	-0.64	1.13
2	UREA I	UREASE BERTHELOT	Biosystems	57	51.80	14.19	7.35	50 mg/dl	-0.24	1.95
3	CREATININE I	JAFFE RATE-BLANKED & COMPENSATED/Alkaline Picrate	Biosystems	187	3.48	9.25	0.32	3.1 mg/dl	-1.18	0.05
4	T.BILIRUBIN I	DIAZONIUM SALT (Colorimetric) /JENDRASSIK	Biosystems	179	5.55	13.16	0.73	5.3 mg/dl	-0.34	0.11
5	T-PROTEIN I	BIURET - colorimetric	Biosystems	174	4.87	9.94	0.48	5.2 g/dl	0.68	0.07
6	ALBUMIN I	BCG - colorimetric	Biosystems	174	2.97	11.06	0.33	3.1 g/dl	0.40	0.05
7	URIC ACID I	ENZYMATIC / URICASE Colorimetric	Biosystems	154	4.58	14.95	0.68	4.7 mg/dl	0.18	0.11
8	CHOLESTEROL I	CHOD-PAP	Biosystems	200	97.77	12.68	12.40	85 mg/dl	-1.03	1.75
9	TRIGLYCERIDE I	GPO-PAP / Enzymatic Colorimetric / End Point	Biosystems	166	93.42	17.66	16.50	90 mg/dl	-0.21	2.56

SDI Range	Interpretation
Within -1.0 to +1.0	Excellent.
Between ± 1.0 to ± 2.0	Good.
Between ± 2.0 to ± 3.0	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

Page 1 of 1

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

Contact details:

Email: clinqc@cmcvellore.ac.in

Contact Number: 0416-2283102



Dr. Pamela Christudoss
 CMC EQAS Co-Ordinator
 Christian Medical College, Vellore

***** End of Report *****