

LAB MONTHLY SUMMARY

Lab Name **SRI KUMARAN HOSPITAL** Lab No **11392**
 Month **April** Year **2020**
 Constituent Group **Chemistry II**

Date of Result Entered : 19/04/2020

Date of Report Published : 06/06/2020

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE I	GOD-POD	Biosystems	206	189.55	8.75	16.58	178 mg/dl	-0.70	2.31
2	UREA I	UREASE BERTHELOT	Biosystems	58	73.63	16.77	12.35	80 mg/dl	0.52	3.24
3	CREATININE I	JAFFE RATE-BLANKED & COMPENSATED/Alkaline Picrate	Biosystems	113	1.35	12.93	0.18	1.2 mg/dl	-0.86	0.03
4	T.BILIRUBIN I	DIAZONIUM SALT (Colorimetric) /JENDRASSIK	Biosystems	164	4.27	14.27	0.61	4.8 mg/dl	0.87	0.10
5	T-PROTEIN I	BIURET - colorimetric	Biosystems	168	4.89	10.30	0.50	5.1 g/dl	0.42	0.08
6	ALBUMIN I	BCG - colorimetric	Biosystems	178	2.97	10.37	0.31	2.6 g/dl	-1.20	0.05
7	URIC ACID I	ENZYMATIC / URICASE Colorimetric	Biosystems	166	8.08	14.55	1.18	9.1 mg/dl	0.87	0.18
8	CHOLESTEROL I	CHOD-PAP	Biosystems	188	94.21	12.27	11.56	88 mg/dl	-0.54	1.69
9	TRIGLYCERIDE I	GPO-PAP / Enzymatic Colorimetric / End Point	Biosystems	173	182.82	12.75	23.30	185 mg/dl	0.09	3.54

SDI Range	Interpretation
Within -1.0 to +1.0	Excellent.
Between ±1.0 to ±2.0	Good.
Between ±2.0 to ±3.0	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Page 1 of 1

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

Contact details:

Email:clinqc@cmcvellore.ac.in

Contact Number: 0416-2283102

Pamela Christudoss

Dr. Pamela Christudoss
 CMC EQAS Co-Ordinator
 Christian Medical College, Vellore

***** End of Report *****