

## LAB MONTHLY SUMMARY

Lab Name **AB DIAGNOSTICS**Lab No **11537**Month **May**Year **2020**Constituent Group **Chemistry II**Date of Result Entered : **19/06/2020**Date of Report Published : **04/08/2020**

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE I	GOD-POD	Mindray	104	106.28	10.92	11.61	89.2 mg/dl	-1.47	2.28
2	UREA I	NED DYE	Any Analyser	80	31.75	15.49	4.92	32.1 mg/dl	0.07	1.10
3	CREATININE I	JAFFE RATE-BLANKED & COMPENSATED/Alkaline Picrate	Mindray	71	4.52	12.26	0.55	4.4 mg/dl	-0.22	0.13
4	T.BILIRUBIN I	DIAZONIUM SALT ( Colorimetric) /JENDRASSIK	Mindray	65	2.71	13.19	0.36	3.3 mg/dl	1.65	0.09
5	T-PROTEIN I	BIURET - colorimetric	Mindray	77	5.21	8.59	0.45	4.7 g/dl	-1.14	0.10
6	ALBUMIN I	BCG - colorimetric	Mindray	81	3.28	9.03	0.30	3.8 g/dl	1.76	0.07
7	URIC ACID I	ENZYMATIC / URICASE Colorimetric	Mindray	87	6.29	13.42	0.84	4.5 mg/dl	-2.12	0.18
8	CHOLESTEROL I	CHOD-PAP	Mindray	94	108.35	8.80	9.53	104.8 mg/dl	-0.37	1.97
9	TRIGLYCERIDE I	GPO-PAP / Enzymatic Colorimetric / End Point	Mindray	80	238.99	10.44	24.94	254.8 mg/dl	0.63	5.58

SDI Range	Interpretation
Within -1.0 to +1.0	Excellent.
Between $\pm 1.0$ to $\pm 2.0$	Good.
Between $\pm 2.0$ to $\pm 3.0$	Accept with caution. Warning Signal.
Beyond $\pm 3.0$	Unacceptable performance. Action Signal.

Page 1 of 1

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

Contact details:

Email:clinqc@cmcvellore.ac.in

Contact Number: 0416-2283102

*Pamela Christudoss*

Dr. Pamela Christudoss

CMC EQAS Co-Ordinator

Christian Medical College, Vellore

\*\*\*\*\* End of Report \*\*\*\*\*