

## Ortho Clinical Diagnostics

403, Leela Business Park,  
Andheri Kurla Road, Andheri  
East, Mumbai - 400059

T : +91 22 6787 9300

F : +91 22 6787 9333

### CALIBRATION CERTIFICATE

CERTIFICATE NO : OCD/PRIMA DIAGNOSTICS/01

CALIBRATION DATE :24/12/2019

CALIBRATED FOR : PRIMA DIAGNOSTICS

CALIBRATION DUE :23/06/2020

LOCATION : MYSORE

CALIBRATED AT : AT SITE

SPECIFICATION OF INSTRUMENT :

INSTRUMENT	MODEL	SERIAL NO
VITROS CHEMISTRY SYSTEM	VITROS 250 J NUMBER	27000023 J27000023

*Specification of Software:*

SOFTWARE NAME	SOFTWARE VERSION	UPDATED ON
DOS Rmx os	V9.7	September 2019

*The Reference of Calibration:* The performance/Adjustments of various subsystems has been calibrated/tested by trained site engineer with pre-designed /calibrated tools provided for the particular subsystems by the company.

#### PM Procedure Checklists:

##### Subsystem Name

##### Adjusted/Verified

##### 1. Sample Supply

\* Hook Drive Belt



##### 2. Slide Supply 1, 2 & slide transport

Mechanical Adjustment:

- \* Dispense Blade Centering Adjustment
- \* Dispense Blade Drive Belt Tension
- \* Plunger Centering Alignment

Software Adjustment:

- \* Dispense To Metering Position



##### 3. Sample Metering

Mechanical Adjustment:

- \* Truck Sensor



- \* Truck Guide Button
- \* Metering Truck

Ortho Clinical Diagnostics India Private Limited

Registered Office: 403, Leela Business Park, Andheri Kurla Road, Andheri East, Mumbai - 400059 | T: +91 22 6787 9300 | F: +91 22 6787 9333 | CIN: U51397MH2015FTC262650

## Ortho Clinical Diagnostics

403, Leela Business Park,  
Andheri Kurla Road, Andheri  
East, Mumbai - 400059

T: +91 22 6787 9300

F: +91 22 6787 9333

### Software Adjustment:

- \* Tip Pickup X,Y and Height
- \* Sample Tray at Sample Position –Auto Sample Tray
- \* Proboscis Centering and Sample Cup Depth
- \* Mixing Cup X,Y and its Depth
- \* Diluents Tray
- \* Rate / Colorimetric Metering Position
- \* Potentiometric Metering Position
- \* Proboscis Compression
- \* Wetness Detector



### 6. Slide Incubator

#### Software Adjustment:

- \* Incubator Disk Sizing
- \* Incubator Disk Stopping
- \* PM Slide Insertion Depth
- \* Slide Read Position (RT/CM Depth & Read Sync)
- \* Potentiometric Eject Blade Depth



### 7. Reflectometer

- \* IRIS Lamp Output Voltage
- \* Correction Factor




### 9. Master Computer

- \* Touch Screen Calibration
- \* System Full Backup



### CALIBRATED BY



Mr.Karthik(Service Engineer)

### NOTE:

1. This certificate refers only to the particular item submitted for calibration.
2. The calibration result reported in the certificate is valid at the time of and under the stated condition of the measurement.

Ortho Clinical Diagnostics India Private Limited

Registered Office: 403, Leela Business Park, Andheri Kurla Road, Andheri East, Mumbai -  
400059 | T: +91 22 6787 9300 | F: +91 22 6787 9333 | CIN: U51397MH2015FTC262650