


CERTIFICATE OF CALIBRATION

Ref. No.: SJH 169
 Applicant: DHARMAGIRI ST. JOSEPH'S HOSPITAL
 Device Calibrated: FULLY AUTO ANALYSER
 Make/Model: VITROS SYSTEM
 Sl. No: ..

Condition at receipt: Functional
 Calibrated at : Onsite
 Date of Calibration : 3/06/2020
 Calibration due date : 2/06/2021

Certified that the above Instrument has been calibrated by trained technical personnel of this centre using a calibration system traceable to National Standards. The calibration details attached with the certificate are authentic quantitative analysis of the instrument's credibility on the date of calibration and relate only to the Instrument calibrated.

Date of Issue : 4/06/2020


 Approved Signatory



Cyrix Healthcare Pvt. Ltd.,
 30/641 B Petta Junction, Poonithura, Kochi-38.
 Ph : 97455 55073
 Email: sales.kl@cyrix.in
 service.kl@cyrix.in



www.cyrixhealthcare.com

ISO Certified Company | AERB Approved QA Agency | NABL Accredited Calibration Agency

CERTIFICATE OF CALIBRATION

Customer Name & Address	DHARMAGIRI ST. JOSEPH'S HOSPITAL, AGASTIAMUZZHI MUKKAM, KOZHIKKODE KERALA - 673602	Certificate No.	SJH 169
		Date of Calibration	3/06/2020
		Valid Up to	2/06/2021
		Date of Issue	4/06/2020

Instruments Details			
Instrument Name	FULLY AUTO ANALYSER	S No.	..
Make/Model No.	VITROS SYSTEM	Location	LABORATORY
ID No.	SJH/LAB/FULLY/01	Visual Inspection	OK

Details of Reference Standards & Major Equipment's Used		
Equipment Name	ELECTRICAL SAFETY ANALYZER	THERMO HYGROMETER
Make	RIGEL MEDICAL	...
Model/SR No.	288+/05L-0407	HTC-2
Calibration Validity	14/06/2020	23/09/2020
Calibration By	ARAGON CALIBRATION	STIC

Environmental Condition			
Temperature	27°C	Relative Humidity	54%

Calibration Results	
ELECTRICAL SAFETY REPORT ATTACHED	

REMARKS: -

- ❖ This result of calibration refers to only to the particular items submitted for calibration.
- ❖ The above result is valid at the time of and under the stated conditions measurement.
- ❖ Calibration certificate issued for weight & measure parameters like mass, weighting balance, volumetric equipment, measuring scales/tapes etc. Are scientific purpose only and should not used for commercial.

Calibrated By

(SC)



Approved By

(QM)



Cyrix Healthcare Pvt. Ltd.,
30/641 B Petta Junction, Poonithura, Kochi-38.
Ph: 97455 55073
Email: sales.kl@cyrix.in
service.kl@cyrix.in

CYRIXTM
HEALTH CARE PVT LTD

www.cyrixhealthcare.com

CERTIFICATE OF CALIBRATION

ELECTRICAL SAFETY TEST VALUES

TEST NO.	NAME OF THE TEST	READING	HIGH LIMIT	LOW LIMIT	UNIT	STATUS
1	MAIN VOLTAGE -LIVE TO NEUTRAL	229	250	225	V	P
2	MAIN VOLTAGE -LIVE TO EARTH	228	240	220	V	P
3	MAIN VOLTAGE-NEUTRAL TO EARTH	0.6	3	0	V	P
4	EQUIPMENT CURRENT	0.2	16	0.1	mA	P
5	PROTECTIVE EARTH RESISTANCE	0.7	3	0	Ω	P
6	EARTH LEAKAGE CURRENT - NC	192	250	50	μ A	P
7	EARTH LEAKAGE CURRENT - OC	291	500	250	μ A	P
8	ENCLOSURE LEAKAGE CURRENT -NC	0.2	10	0	μ A	P
9	ENCLOSURE LEAKAGE CURRENT -OC	0.7	10	0	μ A	P
10	INSULATION RESITANCE					
	1.MAINS-PROTECTIVE EARTH	OL	100	0	M Ω	P
	2.APPLIED PART TO PROTECTIVE EARTH	NA	100	0	M Ω	P
	3.MAINS TO APPLIED PART	NA	100	0	M Ω	P
	4.MAINS TO NEUTRAL	OL	100	0	M Ω	P
	5.APPLIED PART TO NEUTRAL	NA	100	0	M Ω	P

* NOTE (NC-NORMAL CONDITION, OC-OPEN CONDTION)

Service Engineer

Quality Manager

Cyrix Healthcare Pvt. Ltd.,
 30/641 B Petta Junction, Poonithura, Kochi-38.
 Ph: 97455 55073
 Email: sales.kl@cyrix.in
 service.kl@cyrix.in

CYRIXTM
 HEALTH CARE PVT LTD

www.cyrixhealthcare.com

CERTIFICATE OF CALIBRATION


Ref. No.: SJH 173
Applicant: DHARMAGIRI ST. JOSEPH'S HOSPITAL
Device Calibrated: PROTEIN ANALYSER
Make/Model: MISPA I
Sl. No: 21151620837

Condition at receipt: Functional
Calibrated at : Onsite
Date of Calibration : 3/06/2020
Calibration due date : 2/06/2021

Certified that the above Instrument has been calibrated by trained technical personnel of this centre using a calibration system traceable to National Standards. The calibration details attached with the certificate are authentic quantitative analysis of the instrument's credibility on the date of calibration and relate only to the Instrument calibrated.

Date of Issue :

4/06/2020


Approved Signatory



Cyrix Healthcare Pvt. Ltd.,
30/641 B Petta Junction, Poonithura, Kochi-38.
Ph : 97455 55073
Email: sales.kl@cyrix.in
service.kl@cyrix.in

CYRIXTM
HEALTHCARE PVT LTD

www.cyrixhealthcare.com

ISO Certified Company

AERB Approved
QA Agency

NABL Accredited
Calibration Agency

CERTIFICATE OF CALIBRATION

Customer Name & Address	DHARMAGIRI ST. JOSEPH'S HOSPITAL, AGASTIAMUZHI MUKKAM, KOZHIKKODE KERALA - 673602	Certificate No.	SJH 173
		Date of Calibration	3/06/2020
		Valid Up to	2/06/2021
		Date of Issue	4/06/2020

Instruments Details

Instrument Name	PROTEIN ANALYSER	S No.	21151620837
Make/Model No.	MISPA I	Location	LABORATORY
ID No.	SJH/LAB/MISPA/01	Visual Inspection	OK

Details of Reference Standards & Major Equipment's Used

Equipment Name	ELECTRICAL SAFETY ANALYZER	THERMO HYGROMETER
Make	RIGEL MEDICAL	...
Model/SR No.	288+/05L-0407	HTC-2
Calibration Validity	14/06/2020	23/09/2020
Calibration By	ARAGON CALIBRATION	STIC

Environmental Condition

Temperature	27°C	Relative Humidity	54%
-------------	------	-------------------	-----

Calibration Results

ELECTRICAL SAFETY REPORT ATTACHED

REMARKS: -

- ❖ This result of calibration refers to only to the particular items submitted for calibration.
- ❖ The above result is valid at the time of and under the stated conditions measurement.
- ❖ Calibration certificate issued for weight & measure parameters like mass, weighting balance, volumetric equipment, measuring scales/tapes etc. Are scientific purpose only and should not used for commercial.

Calibrated By
(SE)

Approved By
(QM)



Cyrix Healthcare Pvt. Ltd.,
30/641 B Petta Junction, Poonithura, Kochi-38.
Ph: 97455 55073
Email: sales.kl@cyrix.in
service.kl@cyrix.in

CYRIXTM
HEALTH CARE PVT LTD

www.cyrixhealthcare.com

CERTIFICATE OF CALIBRATION

ELECTRICAL SAFETY TEST VALUES

TEST NO.	NAME OF THE TEST	READING	HIGH LIMIT	LOW LIMIT	UNIT	STATUS
1	MAIN VOLTAGE -LIVE TO NEUTRAL	229	250	225	V	P
2	MAIN VOLTAGE -LIVE TO EARTH	228	240	220	V	P
3	MAIN VOLTAGE-NEUTRAL TO EARTH	0.6	3	0	V	P
4	EQUIPMENT CURRENT	0.2	16	0.1	mA	P
5	PROTECTIVE EARTH RESISTANCE	0.7	3	0	Ω	P
6	EARTH LEAKAGE CURRENT - NC	192	250	50	μ A	P
7	EARTH LEAKAGE CURRENT - OC	291	500	250	μ A	P
8	ENCLOSURE LEAKAGE CURRENT -NC	0.2	10	0	μ A	P
9	ENCLOSURE LEAKAGE CURRENT -OC	0.7	10	0	μ A	P
10	INSULATION RESITANCE					
	1.MAINS-PROTECTIVE EARTH	OL	100	0	M Ω	P
	2.APPLIED PART TO PROTECTIVE EARTH	NA	100	0	M Ω	P
	3.MAINS TO APPLIED PART	NA	100	0	M Ω	P
	4.MAINS TO NEUTRAL	OL	100	0	M Ω	P
	5.APPLIED PART TO NEUTRAL	NA	100	0	M Ω	P

* NOTE (NC-NORMAL CONDITION, OC-OPEN CONDTION)

Service Engineer

Quality Manager

Cyrix Healthcare Pvt. Ltd.,
 30/641 B Petta Junction, Poonithura, Kochi-38.
 Ph: 97455 55073
 Email: sales.kl@cyrix.in
 service.kl@cyrix.in

CYRIXTM
 HEALTH CARE PVT LTD

www.cyrixhealthcare.com