

REMI

Eq. Name: CENTRIFUGE
Make: REMI
Model: C 852
EQ. ID: JCL-LAB
Service Date: 02.01.20
Signature: S. B.
Performance Analysis Date:
Performance Analysis by:
Ace Medical Solutions
(India) Pvt. Ltd. Madurai-625 002
ISO 9001:2015 Certified
www.acemedical.in email: sales@acemedical.in
For Service: 94454 74620
For Calibration: 987 000 8750



Ace Medical Solutions (India) Pvt. Ltd.

An ISO 9001-2015 Certified Company

No : 3/751, 1st Cross Street, Surveyor Colony, K.Pudur, Madurai - 625 007.

Phone : 0452 - 2561691 Mobile : +91 94434 74610



GSTIN. 33AAHCA8053H1ZR • DL No. MDU/4205/20B,MDU/3971/21B • CIN : U33110TN2009PTC072140

CALIBRATION CERTIFICATE

Customer Name & Address:	M/S. Jeyam Clinical Lab, K.Pudur, Ayyappan Kovil, Madurai.	Certificate No:	AMS/LAB/2887250
		Issuing Date:	03/01/2020
		Page No:	1 of 1

Details of Device under Calibration (DUC):

DUC:	Centrifuge	Id. No:	JCL/AYKL/CEN/001
Make:	Remi	Department:	Laboratory
Model:	C852	Calibrated On:	02/01/2020
Sl. No:	ZDKN 32771	Calibration Due Suggested:	01/01/2021
Type:	Manual	DUC Condition on Receipt:	OK
Range & Least Count:	0 - 3500 / LC - 1		

Traceability Details:

Sl.No.	Test Equipment	Make	Certificate Number	Valid Up to
1	Digital Tachometer	Lutron / DT-2234C	FSMS/CL/05/19-07/203-01	30/07/2020
2	Thermo Hygrometer	HTC	QC/MD/20190218	10/02/2020

Room Temperature: 25°C±2

Room Humidity: 30-60%

(At the time of Calibration)

Test Results:

SL.No.	DUC Reading (RPM)	Standard Reading (RPM)	Correction (RPM)
1	Position 1	1548	0
2	Position 2	1850	0
3	Position 3	2570	0
4	Position 4	2936	0
5	Position 5	3241	0

Allowable Tolerance: Set by Customer

Note:

1. Calibration Points were selected as per customer request.
2. Each reading were taken by Repeatability Test (i.e.) each reading in a row is a mean of 10 readings.
3. Results Reported are valid at the time of and under the stated conditions of measurements.

Calibrated By:

(Calibration Engineer)

B. B. S.



Approved by:
(Technical Manager)

B. B. S.