



Navess
ON/OFF
T/T

4.9°C



Haier



POWER SAVINGS GUIDE

ELECTRICITY CONSUMPTION
212*
UNITS PER YEAR

Label Period: 01st Jan 2017 - 31st Dec 2018

Appliance	Refrigerator
Brand	Haier
Model / Year	HRD-1803 / 2017
Type	Direct Cool
Gross Volume	180 litre
Storage Volume	182 litre

*Under test conditions, when tested in accordance with approved standards. Actual electricity consumption will depend on how the appliance is used.

Ace Medical Solutions (India) Pvt. Ltd.

An ISO 9001-2015 Certified Company

No : 3/751, 1st Cross Street, Surveyor Colony, K.Pudur, Madurai - 625 007.

Phone : 0452 - 2561691 Mobile : +91 94434 74610



GSTIN: 33AAHCA8053H1ZR • DL No. MDU/4205/20B,MDU/3971/21B • CIN : U33110TN2009PTC072140

CALIBRATION CERTIFICATE

Customer Name & Address:	M/S. Jeyam Clinical Lab, K.Pudur, Ayyappan Kovil, Madurai.	Certificate No:	AMS/LAB/2887257
		Issuing Date:	03/01/2020
		Page No:	1 of 1

Details of Device under Calibration (DUC):

DUC:	Refrigerator – Reagent	Id. No:	JCL/AYKL/RREF/001
Make:	Haier	Department:	Laboratory
Model:	HRD1903	Calibrated On:	02/01/2020
Sl. No:	---	Calibration Due Suggested:	01/01/2021
Type:	Drug / Reagent	DUC Condition on Receipt:	OK
Range & Least Count:	0 – 8°C / LC 1°C		

Traceability Details:

Sl.No.	Test Equipment	Make	Certificate Number	Valid Up to
1	RTD Thermometer	Center / RTD 375	QC/MD/20191121	09/02/2020
2	Thermo Hygrometer	HTC	QC/MD/20190218	10/02/2020

Room Temperature: 25°C±2 Room Humidity: 30-60% (At the time of Calibration)

Test Results:

SL.No.	DUC Reading (°C)	Standard Reading (°C)	Correction (°C)
1	5.0	4.98	-0.02
2	4.0	4.03	+0.03

Allowable Tolerance: ±1°C

Note:

1. Calibration Points were selected as per customer request.
2. Each reading were taken by Repeatability Test (i.e.) each reading in a row is a mean of 10 readings.
3. Results Reported are valid at the time of and under the stated conditions of measurements.

Calibrated By:

(Calibration Engineer)

B. S. S.



Approved by:

(Technical Manager)

B. S. S.