



CALIBRATION REPORT

Details of the Unit Under Calibration:

Calibration Certificate No: SBS/CL/20-21/01899

Page: 1 of 2

| | | | |
|--------------------------|-------------------------|-----------------------------|----------------------------------|
| Customer Name: | SRI KUMARAN HOSPITAL | Customer Address: | VIRALIMALAI ROAD, MANAPPARAI. |
| Location: | LAB | Serial No: | 041600508 |
| Description: | PIPETTE | Date of Calibration: | 27-10-2020 |
| Make & Model: | BIOSYSTEMS & HANDYPETTE | Date of Due: | 26-10-2020 |
| Range: | 100-1000 μ L | | VARIABLE |

Traceability Details:

| S.No | Standard Used | Make | Model | Serial No | Validity |
|------|-------------------------------|---------|--------|-----------|-------------|
| 1. | Electronic Semi Micro Balance | Citizen | Cx265N | 477904 | 08-Apr-2021 |

Environmental Conditions and Standard Operating Procedure Details:

| <i>ENVIRONMENTAL CONDITIONS AND STANDARD OPERATING PROCEDURE DETAILS</i> | | | |
|--|------|--------------------------|--------|
| Temperature | 24°C | Calibration Done at | ONSITE |
| Humidity(40-70% RH) | 55 % | Calibration Procedure No | -- |

REMARKS:

Test Report : Attached

Calibrated by:

M.BALAJI

(Calibration Engineer)

Approved by:

C.SHANMUGARAJ

(Technical Manager)





Result of Calibration: SBS/TR/20-21/01899

Page: 2 of 2

| TEST REPORT | | | | |
|---------------|----------------|-----------------|-----------|--------|
| Units | Nominal Values | Measured Values | Tolerance | Result |
| μL | 100 | 100.2 | 0.5 % | PASS |
| μL | 500 | 500.0 | 0.5 % | PASS |
| μL | 1000 | 1000.0 | 0.5 % | PASS |



CALIBRATION RESULT:

- Corrective maintenance required
- Removed from use
- Acceptable for use

Calibration Engineer Sign.

Date: 27.10.2020