

CALIBRATION REPORT

Details of the Unit Under Calibration:

Calibration Certificate No: SBS/CL/20-21/01902

Page: 1 of 2

Customer Name:	SRI KUMARAN HOSPITAL	Customer Address:	VIRALIMALAI ROAD, MANAPPARAI.
Location:	LAB	Serial No:	091501183
Description:	PIPETTE	Date of Calibration:	27-10-2020
Make & Model:	BIOSYSTEMS & HANDYPETTE	Date of Due:	26-10-2020
Range:	10µL		FIXED

Traceability Details:

S.No	Standard Used	Make	Model	Serial No	Validity
1.	Electronic Semi Micro Balance	Citizen	Cx265N	477904	08-Apr-2021

Environmental Conditions and Standard Operating Procedure Details:

ENVIRONMENTAL CONDITIONS AND STANDARD OPERATING PROCEDURE DETAILS			
Temperature	24°C	Calibration Done at	ONSITE
Humidity(40-70% RH)	55 %	Calibration Procedure No	--

REMARKS:

Test Report : Attached

Calibrated by:



M.BALAJI

(Calibration Engineer)

Approved by:




C.SHANMUGARAJ
(Technical Manager)



Result of Calibration: SBS/TR/20-21/01902

Page: 2 of 2

TEST REPORT				
Units	Nominal Values	Measured Values	Tolerance	Result
μL	10	10.1	0.5 %	PASS



CALIBRATION RESULT:

- Corrective maintenance required
- Removed from use
- Acceptable for use

Calibration Engineer Sign: 

Date: 27.10.2020