



CALIBRATION REPORT

Details of the Unit Under Calibration:

Calibration Certificate No: SBS/CL/20-21/01901

Page: 1 of 2

Customer Name:	SRI KUMARAN HOSPITAL	Customer Address:	VIRALIMALAI ROAD, MANAPPARAI.
Location:	LAB	Serial No:	190411026
Description:	PIPETTE	Date of Calibration:	27-10-2020
Make & Model:	BIOSYSTEMS & HANDYPETTE	Date of Due:	26-10-2020
Range:	100-1000 μ L		VARIABLE

Traceability Details:

S.No	Standard Used	Make	Model	Serial No	Validity
1.	Electronic Semi Micro Balance	Citizen	Cx265N	477904	08-Apr-2021

Environmental Conditions and Standard Operating Procedure Details:

ENVIRONMENTAL CONDITIONS AND STANDARD OPERATING PROCEDURE DETAILS			
Temperature	24°C	Calibration Done at	ONSITE
Humidity(40-70% RH)	55 %	Calibration Procedure No	--

REMARKS:

Test Report : Attached

Calibrated by:

M.BALAJI

(Calibration Engineer)

Approved by:

C.SHANMUGARAJ

(Technical Manager)





Result of Calibration: SBS/TR/20-21/01901

Page: 2 of 2

TEST REPORT				
Units	Nominal Values	Measured Values	Tolerance	Result
μL	100	100.1	0.5 %	PASS
μL	500	500.0	0.5 %	PASS
μL	1000	1000.0	0.5 %	PASS



CALIBRATION RESULT:

- Corrective maintenance required
- Removed from use
- Acceptable for use

Calibration Engineer Sign 

Date: 27.10.2020