

## LAB MONTHLY SUMMARY

Lab Name **SARDANA LABS**Lab No **2314**Month **September**Year **2020**Constituent Group **Chemistry II**Date of Result Entered : **20/09/2020**Date of Report Published : **07/10/2020**

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE I	GOD-POD	Any Analyser	838	223.99	10.76	24.09	192 mg/dl	-1.33	1.66
2	UREA I	UREASE UV / GLDH	Any Analyser	476	127.21	14.07	17.90	128 mg/dl	0.04	1.64
3	CREATININE I	JAFFE-END POINT	Any Analyser	340	1.40	12.32	0.17	1.3 mg/dl	-0.58	0.02
4	T.BILIRUBIN I	DIAZONIUM SALT (Colorimetric) /JENDRASSIK	Any Analyser	588	5.24	14.11	0.74	5 mg/dl	-0.32	0.06
5	ALBUMIN I	BCG - colorimetric	Any Analyser	603	3.23	9.57	0.31	3.9 g/dl	2.17	0.03
6	URIC ACID I	ENZYMATIC / URICASE Colorimetric	Any Analyser	645	7.12	15.23	1.08	6.1 mg/dl	-0.94	0.09
7	CHOLESTEROL I	CHOD-PAP	Any Analyser	726	108.64	11.67	12.68	102 mg/dl	-0.52	0.94

SDI Range	Interpretation
Within -1.0 to +1.0	Excellent.
Between $\pm 1.0$ to $\pm 2.0$	Good.
Between $\pm 2.0$ to $\pm 3.0$	Accept with caution. Warning Signal.
Beyond $\pm 3.0$	Unacceptable performance. Action Signal.

Page 1 of 1

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

Contact details:

Email: [clinqc@cmcvellore.ac.in](mailto:clinqc@cmcvellore.ac.in)

Contact Number: 0416-2283102

*Pamela Christudoss*

Dr. Pamela Christudoss  
 CMC EQAS Co-Ordinator  
 Christian Medical College, Vellore

\*\*\*\*\* End of Report \*\*\*\*\*