

Date: 21/07/2020

To,
Dr. Satav's Pathology Laboratory,
Pune.

Kind Attn : Mr. Rameshwar,

Calibration Certificate

We hereby certify that the Bio - Rad D - 10 system Sr. No. DJ4F013423 Installed in your lab is calibrated today as under & the performance was found satisfactory.

Pump flow calibration @ 1.5 ml / min (Flow checked by using volumetric variant sample cups with visible observation)

Thermo module temperature calibration @ 35^o (Temp. Checked by using D - 10 internal inbuilt temp. Sensor)

Detector log value @ 0.020 (Voltage verified by using Multimeter)

The next calibration is due on: 20/07/2021

For,
Anand Chemiceutics


Authorised Signatory