

LAB MONTHLY SUMMARY

Lab Name DR.SATAV'S PATHOLOGY LABORATORY Lab No 2438
 Month April Year 2020
 Constituent Group HbA1c

Date of Result Entered : 19/04/2020

Date of Report Published : 06/06/2020

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	HbA1c	HPLC-ION EXCHANGE	Biorad (D10 / D100 / Variant Turbo)	448	5.46	7.56	0.41	5.4 %	-0.15	0.04

SDI Range	Interpretation
Within -1.0 to +1.0	Excellent.
Between ±1.0 to ±2.0	Good.
Between ±2.0 to ±3.0	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Page 1 of 1

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

Contact details:

Email:clinqc@cmcvellore.ac.in

Contact Number: 0416-2283102

Pamela Christudoss

Dr. Pamela Christudoss

CMC EQAS Co-Ordinator

Christian Medical College, Vellore

***** End of Report *****