

**Biochemistry QC Chart**

**16/10/2020**


This is certify that the Biochemistry Analyzer model Abachem 200 Plus with serial No:ATCD6680818AHC is calibrated with known value controls and results shown below.

Calibration Date:16/10/2020 Next Calibration Date:15/04/2021

Customer Name: Medicity Referral Laboratory  
Thaliparamba

<b>RANDOX Lot no: 1310UN Ex.Date: 2021-05-28</b>			
<b>S.No</b>	<b>PARAMETER</b>	<b>UNIT</b>	<b>LEVEL 2</b>
1	Albumin	g/dL	4.10
2	Alkaline Phosphates	U/L	282
3	ALT/SGPT	U/L	33
4	Amylase	U/l	73
5	AST/SGOT	U/L	33
6	Bilirubin (Direct)	mg/dl	1.12
7	Bilirubin(Total)	mg/dl	1.5
8	Calcium	mg/dl	8.8
9	Chloride	mEq/L	97.8
10	Cholesterol(HDL)	mg/dl	61.5
11	Cholesterol (Total)	mg/dl	192
12	Creatinine	mg/dl	1.29
13	Glucose	mg/dl	125
14	Phosphorus	mg/dl	7.12
15	Triglycerides	mg/dl	93.0
16	Urea	mg/dl	44.5
17	Urea Nitrogen(BUN)	mg/dl	45.7
18	Uric Acid	mg/dl	5.6
19	Potassium	mmol/L	3.89
20	Sodium	mmol/L	142

For,



*[Handwritten Signature]*

Abaxis Healthcare India Pvt Ltd

**Abaxis Healthcare India Pvt. Ltd**

#134, Gandhi Nagar Main, Cheroor Post, Thrissur 680008, Kerala, India

Phone.+919961855000, 0487-2323537

Mail ID. [abaxishealthcare@gmail.com](mailto:abaxishealthcare@gmail.com), [sales@abaxishealthcare.com](mailto:sales@abaxishealthcare.com)

web: [www.abaxishealthcare.com](http://www.abaxishealthcare.com)