

MINDRAY – BS240 Auto Bio Chemistry Analyser

Customer Address: **Medcity Referral Laboratoy**
Opp.Bustand, Bell Square Complex
(Ground Floor), Near Badariya Hotel
Thaliparampa.

Contact person: Mr. Rajeesh
Telephone No: 07012518717
Serial no: YX-79001079
Calibrated date: 09/10/2020

Calibration Certificate

CALIBRATION

The Analyser has been calibrated with Mindray Multi Sera Calibrator for Glucose, Creatinine and Urea, T.Bilirubin, D.Bilirubin, ALT, AST, ALP, TP, Uric Acid, Cholesterol and ALB.

Calibrator Manufacture : Mindray
Calibrator name : Multi Sera Calibrator
Lot no. : 150119002
Expiry date : 31-05-2021

Carried out Multi Sera Calibration found values are within specified range.

REMARKS

The system meets all criteria outlined in this Calibration Certificate Documents. All the required parameters have been checked and instrument is ready for regular use.

Performer : **Akhil E.V**
Designation : Application specialist
Date : 10/10/2020



Reviewer : **Muthu vigneshwar . K.P**
Designation : Application Manager

Muthu Vigneshwar . K.P
10/10/2020