



ATHENESE-DX

Early diagnosis for better life

CALIBRATION CERTIFICATE

Certificate No. : ELISA READER -ADX110-004/2020-21
Date of Calibration : 09.05.2020
Calibration Due on : 08.05.2021
Page No. : 01/01

DETAILS OF UNIT UNDER CALIBRATION

Model Name: ALTA ELISA READER(ADX-110)	Sr. No.: RT1060118ADX
--	-----------------------

OPTICAL FILM READING CALIBRATION

Wavelength	Film type & No.	Range	Absorbance Obtained
405 nm	SP - 209	0.331 - 0.349	0.336
	DP - 210	0.661 - 0.697	0.681
	TP - 211	0.970 - 1.020	0.997
	2DP - 210	1.302 - 1.388	1.347
	3DP - 210	1.942 - 2.068	1.983
	4DP - 210	2.554 - 2.702	2.641
450 nm	SP - 209	0.312 - 0.328	0.319
	DP - 210	0.608 - 0.640	0.632
	TP - 211	0.895 - 0.941	0.907
	2DP - 210	1.197 - 1.277	1.209
	3DP - 210	1.790 - 1.908	1.842
	4DP - 210	2.364 - 2.510	2.478
492 nm	SP - 209	0.335 - 0.353	0.339
	DP - 210	0.672 - 0.708	0.682
	TP - 211	0.998 - 1.048	1.031
	2DP - 210	1.323 - 1.411	1.356
	3DP - 210	1.970 - 2.098	2.063
	4DP - 210	2.572 - 2.732	2.673
630 nm	SP - 209	0.328 - 0.346	0.340
	DP - 210	0.674 - 0.710	0.683
	TP - 211	1.034 - 1.086	1.062
	2DP - 210	1.328 - 1.416	1.388
	3DP - 210	1.984 - 2.114	2.065
	4DP - 210	2.588 - 2.748	2.610

ALTA Elisa Reader ADX110 Sr. No. RT1060118ADX has been calibrated for above parameter and found to be within its acceptance Criteria.

Calibrated By:

[Signature]
A. SRIKRISHNAN



Regd. Office: A. SRIKRISHNAN
ATHENESE-DX PRIVATE LIMITED
 MODULE NO. 407 & 408, 4th FLOOR, TICEL BIO PARK II,
 NO. 5, CSIR ROAD, TARAMANI, CHENNAI - 600113.
 Ph: +91-44-22541131, Email: info@athenesedx.com
 Website: www.athenesedx.com
 CIN NO.U85100TN2012PTC084632

Strategic Partner of



INTEGRITY

COMMITMENT

SERVICE