



CALIBRATION CERTIFICATE

Certificate No.	MES/20-21/DHAA/CEN/D01	Date of Issue : 31/10/2020
		Calibration Done On : 30/10/2020
		Calibration Due On : 29/10/2021

Customer Name	DISTRICT HOSPITAL, ALUVA
---------------	---------------------------------

DETAILS OF DEVICE UNDER CALIBRATION

Equipment Name	Make / Model	Serial No.	Asset ID	Location
Centrifuge	Rotek	08100	68070305	ICTC Lab

DETAILS OF MASTER EQUIPMENT USED

Sl No.	Master Equipment	Make / Model	Serial No.	Reference ID	Cal. Due
1	Digital Tachometer	Equinox / EQ-801B	S916577	ACS/20-21/102671-01	24/09/2021
2	Digital Thermo Hygrometer	Equinox / EQ-308-CTH	20191602	ACS/MC/20-21/2593-01	18/06/2021

The instruments used for calibration are under valid calibration and are traceable to National Standards.

Environmental Conditions	Temperature	27°C	Relative Humidity	55 %
---------------------------------	-------------	------	-------------------	------


PERFORMANCE DATA


Set Reading	Measured Reading
1	1468 rpm
2	2160 rpm
3	2820 rpm
4	3401 rpm
5	3958 rpm

Calibration Status	Pass	✓	Fail	
---------------------------	------	---	------	--

Calibrated By

Approved By


 Prabin KP (Service Engineer)


 Binoj Thomas (Quality Manager)

NB: Results reported are valid at the time of and under the stated conditions of measurements. This report shall not be reproduced without written approval by Medical Engineering & Services.

