



CALIBRATION CERTIFICATE

Certificate No.	MES/20-21/DHAA/REF/D01	Date of Issue : 31/10/2020
		Calibration Done On : 30/10/2020
		Calibration Due On : 29/10/2021

Customer Name	DISTRICT HOSPITAL, ALUVA
---------------	---------------------------------

DETAILS OF DEVICE UNDER CALIBRATION

Equipment Name	Make / Model	Serial No.	Asset ID	Location
Refrigerator	Kelvinator/ D 200 B	-	68078654	ICTC Lab

DETAILS OF MASTER EQUIPMENT USED

SI No.	Master Equipment	Make / Model	Serial No.	Reference ID	Cal. Due
1	Digital Thermometer	Metravi / DTM 900	10055593	ACS/20-21/I02669-01	23/09/2021
2	Digital Thermo Hygrometer	Equinox / EQ-308CTH	20191602	ACS/MC/20-21/2593- 01	18/06/2021

The instruments used for calibration are under valid calibration and are traceable to National Standards.

Environmental Conditions	Temperature	27 °C	Relative Humidity	55 %
---------------------------------	-------------	-------	-------------------	------

PERFORMANCE DATA

Display	Measured	Tolerance
1.3 °C	1.3 °C	± 0.3 °C
2.4 °C	2.4 °C	± 0.3 °C
4.2 °C	4.3 °C	± 0.3 °C
7.2 °C	7.2 °C	± 0.3 °C


Calibration Status	Pass	✓	Fail	
---------------------------	------	---	------	--

Calibrated By

Approved By

Vivek Mahendran (Service Engineer)

Binoj Thomas (Quality Manager)





CALIBRATION CERTIFICATE

Certificate No.	MES/20-21/DHAA/REF/D02	Date of Issue : 31/10/2020
		Calibration Done On : 30/10/2020
		Calibration Due On : 29/10/2021

Customer Name	DISTRICT HOSPITAL, ALUVA
---------------	---------------------------------

DETAILS OF DEVICE UNDER CALIBRATION

Equipment Name	Make / Model	Serial No.	Asset ID	Location
Refrigerator	Haier / HRD – 1902BBR-E	BSO33CE8701T1L8 V0456	80072771	ICTC Lab

DETAILS OF MASTER EQUIPMENT USED

SI No.	Master Equipment	Make / Model	Serial No.	Reference ID	Cal. Due
1	Digital Thermometer	Metravi / DTM 900	10055593	ACS/20-21/I02669-01	23/09/2021
2	Digital Thermo Hygrometer	Equinox / EQ-308CTH	20191602	ACS/MC/20-21/2593-01	18/06/2021

The instruments used for calibration are under valid calibration and are traceable to National Standards.

Environmental Conditions	Temperature	27 °C	Relative Humidity	55 %
---------------------------------	-------------	-------	-------------------	------


PERFORMANCE DATA

Display	Measured	Tolerance
1.3 °C	1.2 °C	± 0.3 °C
2.4 °C	2.4 °C	± 0.3 °C
4.2 °C	4.3 °C	± 0.3 °C
7.2 °C	7.1 °C	± 0.3 °C

Calibration Status	Pass	✓	Fail	
---------------------------	------	---	------	--

Calibrated By

Approved By


 Vivek Mahendran (Service Engineer)


 Binoj Thomas (Quality Manager)

NB: Results reported are valid at the time of and under the stated conditions of measurements. This report shall not be reproduced without written approval by Medical Engineering & Services.

