


# Urea Quality Control Sheet.

An outlier was found in the report in urea which is a Random Error.

The Application Engineer was called and informed.

Internal Quality control was performed.

Calibrator was used and the machine was recalibrated.



This is one of the subsequent LJ charts that are regularly done for Random errors.