

LAB MONTHLY SUMMARY

Lab Name **JOTHI X-RAY AND LABORATORY** Lab No **9661**
 Month **October** Year **2020**
 Constituent Group **HbA1c**

Date of Result Entered : 12/10/2020

Date of Report Published : 09/11/2020

| Sl.No | Analyte | Method / Principle Name | Analyzer Name | No of Participants | DV | Participants | | Your Value | SDI | U |
|-------|---------|-------------------------|---------------------------------------|--------------------|------|--------------|------|------------|------|------|
| | | | | | | CV | SD | | | |
| 1 | HbA1c | HPLC-ION EXCHANGE | Biorad (D10 / D100 / Variant Turbo) | 461 | 6.60 | 6.37 | 0.42 | 7 % | 0.95 | 0.04 |

| SDI Range | Interpretation |
|----------------------|--|
| Within -1.0 to +1.0 | Excellent. |
| Between ±1.0 to ±2.0 | Good. |
| Between ±2.0 to ±3.0 | Accept with caution. Warning Signal. |
| Beyond ±3.0 | Unacceptable performance. Action Signal. |

Page 1 of 1

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

Contact details:

Email: clinqc@cmcvellore.ac.in

Contact Number: 0416-2283102

Pamela Christudoss

Dr. Pamela Christudoss
CMC EQAS Co-Ordinator
Christian Medical College, Vellore

***** End of Report *****