

CERTIFICATE OF CALIBRATION

Ref. No.: SJH 165
Applicant: DHARMAGIRI ST. JOSEPH'S HOSPITAL
Device Calibrated: DIGITAL THERMOMETER
Make/Model: .. REFRIGERATOR - SAMSUNG /
Sl. No: .. RA 23BD TS1 / XTL

Condition at receipt: Functional
Calibrated at : Onsite
Date of Calibration : 3/06/2020
Calibration due date : 2/06/2021

Certified that the above Instrument has been calibrated by trained technical personnel of this centre using a calibration system traceable to National Standards. The calibration details attached with the certificate are authentic quantitative analysis of the instrument's credibility on the date of calibration and relate only to the Instrument calibrated.

Date of Issue : 4/06/2020


Approved Signatory



Cyrix Healthcare Pvt. Ltd.,
30/641 B Petta Junction, Poonithura, Kochi-38.
Ph : 97455 55073
Email: sales.kl@cyrix.in
service.kl@cyrix.in

CYRIXTM
HEALTH CARE PVT LTD

www.cyrixhealthcare.com

ISO Certified Company

AERB Approved
QA Agency

NABL Accredited
Calibration Agency

CERTIFICATE OF CALIBRATION

Customer Name & Address	DHARMAGIRI ST. JOSEPH'S HOSPITAL, AGASTIAMUZZHI MUKKAM, KOZHICKODE KERALA - 673602	Certificate No.	SJH 165
		Date of Calibration	3/06/2020
		Valid Up to	2/06/2021
		Date of Issue	4/06/2020

Instruments Details			
Instrument Name	DIGITAL THERMOMETER	S No.	..
Make/Model No.	..	LOCATION	LABORATORY
BME ID	SJH/LAB/DTM/01	Visual Inspection	OK

Details of Reference Standards & Major Equipment's Used					
No	Equipment Name	Make/Model	Serial No:	Calibration Validity	Calibration By
1	INFRARED THERMOMETER	HTC	2017084051	13/01/2021	PERFECT CALIBRATION
2	DIGITAL THERMOHYGROMETER	-/HTC-2	..	23/09/2020	STIC

Environmental Condition			
Temperature	23.9°C	Relative Humidity	54.4 %

Calibration Results				
SL NO	SET READING(°C)	MEASURED READING (°C)	TOLERANCE	STATUS
1	2	1.6	±2	P
2	4	4.0	±2	P
3	6	6.0	±2	P

REMARKS: -

- ❖ This result of calibration refers to only to the particular items submitted for calibration.
- ❖ The above result is valid at the time of and under the stated conditions measurement.
- ❖ Calibration certificate issued for weight & measure parameters like mass, weighting balance, volumetric equipment, measuring scales/tapes etc. Are scientific purpose only and should not use for commercial.

Calibrated By
(SE)



Approved By
(QM)

SARATH K B



Cyrix Healthcare Pvt. Ltd.,
30/641 B Petta Junction, Poonithura, Kochi-38.
Ph: 97455 55073
Email: sales.kl@cyrix.in
service.kl@cyrix.in

CYRIX™
HEALTH CARE PVT LTD

CERTIFICATE OF CALIBRATION

Ref. No.: SJH 166
Applicant: DHARMAGIRI ST. JOSEPH'S HOSPITAL
Device Calibrated: DIGITAL THERMOMETER
Make/Model: .. REFRIGRATOR. SAMSUNG 10622
Sl. No: .. RA 23 BOTST/1XTL

Condition at receipt: Functional Calibrated at : Onsite
Date of Calibration : 3/06/2020
Calibration due date : 2/06/2021

Certified that the above Instrument has been calibrated by trained technical personnel of this centre using a calibration system traceable to National Standards. The calibration details attached with the certificate are authentic quantitative analysis of the instrument's credibility on the date of calibration and relate only to the Instrument calibrated.

Date of Issue :

4/06/2020

Approved Signatory



Cyrix Healthcare Pvt. Ltd.,
30/641 B Petta Junction, Poonithura, Kochi-38.
Ph : 97455 55073
Email: sales.kl@cyrix.in
service.kl@cyrix.in

CYRIXTM
HEALTH CARE PVT LTD

www.cyrixhealthcare.com

ISO Certified Company

AERB Approved
QA Agency

NABL Accredited
Calibration Agency

CERTIFICATE OF CALIBRATION

Customer Name & Address	DHARMAGIRI ST. JOSEPH'S HOSPITAL, AGASTIAMUZHI MUKKAM, KOZHIKKODE KERALA - 673602	Certificate No.	SJH 166
		Date of Calibration	3/06/2020
		Valid Up to	2/06/2021
		Date of Issue	4/06/2020

Instruments Details

Instrument Name	DIGITAL THERMOMETER	S No.	..
Make/Model No.	..	LOCATION	LABORATORY
BME ID	SJH/LAB/DTM/02	Visual Inspection	OK

Details of Reference Standards & Major Equipment's Used

No	Equipment Name	Make/Model	Serial No:	Calibration Validity	Calibration By
1	INFRARED THERMOMETER	HTC	2017084051	13/01/2021	PERFECT CALIBRATION
2	DIGITAL THERMOHYGROMETER	-/HTC-2	..	23/09/2020	STIC

Environmental Condition

Temperature	23.9°C	Relative Humidity	54.4 %
-------------	--------	-------------------	--------

Calibration Results

SL NO	SET READING(°C)	MEASURED READING (°C)	TOLERANCE	STATUS
1	2	1.6	±2	P
2	4	3.8	±2	P
3	6	6.0	±2	P

REMARKS: -

- ❖ This result of calibration refers to only to the particular items submitted for calibration.
- ❖ The above result is valid at the time of and under the stated conditions measurement.
- ❖ Calibration certificate issued for weight & measure parameters like mass, weighting balance, volumetric equipment, measuring scales/tapes etc. Are scientific purpose only and should not use for commercial.

Calibrated By
(SE)



Approved By
(QM)
SARATH K B

Cyrix Healthcare Pvt. Ltd.,
30/641 B Petta Junction, Poonithura, Kochi-38.
Ph: 97455 55073
Email: sales.kl@cyrix.in
service.kl@cyrix.in

CYRIX™
HEALTH CARE PVT LTD