

# CERTIFICATE OF CALIBRATION

Ref. No.: SJH007  
Applicant : DHARMAGIRI ST JOSEPH'S HOSPITAL.  
Device Calibrated : MICRO PIPETTE  
Make/Model: EPPENDORF  
Sl.No: 117661

Condition at receipt : Functional  
Calibrated at : Onsite  
Date of Calibration : 29/08/2020  
Calibration due date : 28/02/2021

*Certified that the above Instrument has been calibrated by trained technical personnel of this centre using a calibration system traceable to National Standards. The calibration details attached with the certificate are authentic quantitative analysis of the instrument's credibility on the date of calibration and relate only to the Instrument calibrated.*

Date of Issue : 30/08/2020

Approved Signatory



**Cyrix Healthcare Pvt. Ltd.,**  
30/641 B Petta Junction, Poonithura, Kochi-38.  
Ph : 97455 55073  
Email: sales.kl@cyrix.in  
service.kl@cyrix.in

**CYRIX**<sup>TM</sup>  
HEALTH CARE PVT LTD

[www.cyrixhealthcare.com](http://www.cyrixhealthcare.com)

ISO Certified Company

AERB Approved  
QA Agency

NABL Accredited  
Calibration Agency

# CERTIFICATE OF CALIBRATION

CUSTOMER NAME	DHARMAGIRI ST. JOSEPH'S HOSPITAL		
ADDRESS	Koyilandy - Edavanna Rd, Agastiamuzhi, Mukkam, Kerala 673602	DATE OF CALIBRATION	29/08/2020
CERTIFICATE NO:	CHPL/PC/SJH/007	VALID UP TO	28/02/2021
		DATE OF ISSUE	30/08/2020

INSTRUMENTS DETAILS			
INSTRUMENT NAME	MICRO PIPETTE	SI.NO:	117661
MAKE/MODEL.NO:	EPPENDROF	VISUAL INSPECTION	OK

DETAILS OF REFERENCE STANDARDS & MAJOR EQUIPMENTS USED			
EQUIPMENTS NAME	ANALYTICAL BALANCE	THERMO HYGROMETER	
MAKE	WENSAR	HTC-1/-	
MODEL/SI.NO:	MAB182/39736	..	
CALIBRATION VALIDITY	13/02/2021	23/09/2020	
CALIBRATION BY	STIC	STIC	

ENVIRONMENTAL CONDITION			
TEMPERATURE	25.6°C	RELATIVE HUMIDITY	40.1%

CALIBRATION RESULTS							
SL NO	SET READING ( $\mu$ l)	MEASURED READING ( $\mu$ l)					STATUS
		I	II	III	IV	V	
1	200	200.0	200.01	200.01	200.02	200.02	P

**REMARKS: -**

- ❖ This result of calibration refers to only to the particular items submitted for calibration.
- ❖ The above result is valid at the time of and under the stated conditions measurement.
- ❖ Calibration certificate issued for weight & measure parameters like mass, weighting balance, volumetric equipment, measuring scales/tapes etc. Are scientific purpose only and should not use for commercial.

Calibrated By

(SE)

Approved By

(QM)

SARATH K B



**Cyrix Healthcare Pvt. Ltd.,**  
30/641 B Petta Junction, Poonithura, Kochi-38.  
Ph: 97455 55073  
Email: sales.kl@cyrix.in  
service.kl@cyrix.in

**CYRIX**<sup>TM</sup>  
HEALTH CARE PVT LTD

[www.cyrixhealthcare.com](http://www.cyrixhealthcare.com)



# CERTIFICATE OF CALIBRATION

Ref. No.: SJH008  
Applicant : DHARMAGIRI ST JOSEPH'S HOSPITAL.  
Device Calibrated : MICRO PIPETTE  
Make/Model: EPPENDORF  
Sl.No: 115937

Condition at receipt : Functional  
Calibrated at : Onsite  
Date of Calibration : 29/08/2020  
Calibration due date : 28/02/2021

*Certified that the above Instrument has been calibrated by trained technical personnel of this centre using a calibration system traceable to National Standards. The calibration details attached with the certificate are authentic quantitative analysis of the instrument's credibility on the date of calibration and relate only to the Instrument calibrated.*

Date of Issue : 30/08/2020



*[Signature]*  
Approved Signatory

**Cyrix Healthcare Pvt. Ltd.,**  
30/641 B Petta Junction, Poonithura, Kochi-38.  
Ph : 97455 55073  
Email: sales.kl@cyrix.in  
service.kl@cyrix.in

**CYRIX<sup>TM</sup>**  
HEALTH CARE PVT LTD

[www.cyrixhealthcare.com](http://www.cyrixhealthcare.com)

ISO Certified Company

AERB Approved  
QA Agency

NABL Accredited  
Calibration Agency

## CERTIFICATE OF CALIBRATION

CUSTOMER NAME	DHARMAGIRI ST.JOSEPH'S HOSPITAL		
ADDRESS	Koyilandy - Edavanna Rd, Agastiamuzhi, Mukkam, Kerala 673602	DATE OF CALIBRATION	29/08/2020
CERTIFICATE NO:	CHPL/PC/SJH/008	VALID UP TO	28/02/2021
		DATE OF ISSUE	30/08/2020

INSTRUMENTS DETAILS			
INSTRUMENT NAME	MICRO PIPETTE	SI.NO:	115937
MAKE/MODEL.NO:	EPPENDROF	VISUAL INSPECTION	OK

DETAILS OF REFERENCE STANDARDS & MAJOR EQUIPMENTS USED		
EQUIPMENTS NAME	ANALYTICAL BALANCE	THERMO HYGROMETER
MAKE	WENSAR	HTC-1/-
MODEL/SI.NO:	MAB182/39736	..
CALIBRATION VALIDITY	13/02/2021	23/09/2020
CALIBRATION BY	STIC	STIC

ENVIRONMENTAL CONDITION			
TEMPERATURE	25.6°C	RELATIVE HUMIDITY	40.1%

CALIBRATION RESULTS							
SL NO	SET READING ( $\mu$ l)	MEASURED READING ( $\mu$ l)					STATUS
		I	II	III	IV	V	
1	20	21.03	20.03	20.04	20.04	20.02	P

**REMARKS: -**

- ❖ This result of calibration refers to only to the particular items submitted for calibration.
- ❖ The above result is valid at the time of and under the stated conditions measurement.
- ❖ Calibration certificate issued for weight & measure parameters like mass, weighting balance, volumetric equipment, measuring scales/tapes etc. Are scientific purpose only and should not use for commercial.

*[Signature]*  
**Calibrated By**  
 (SE)

*[Signature]*  
**Approved By**  
 (QM)  
 SARATH K B



**Cyrix Healthcare Pvt. Ltd.,**  
 30/641 B Petta Junction, Poonithura, Kochi-38.  
 Ph: 97455 55073  
 Email: sales.kl@cyrix.in  
 service.kl@cyrix.in

**CYRIX™**  
 HEALTH CARE PVT LTD



# CERTIFICATE OF CALIBRATION

Ref. No.: SJH009  
Applicant : DHARMAGIRI ST JOSEPH'S HOSPITAL.  
Device Calibrated : MICRO PIPETTE  
Make/Model: ERBA  
Sl.No: HR62975

Condition at receipt : Functional  
Calibrated at : Onsite  
Date of Calibration : 29/08/2020  
Calibration due date : 28/02/2021

*Certified that the above Instrument has been calibrated by trained technical personnel of this centre using a calibration system traceable to National Standards. The calibration details attached with the certificate are authentic quantitative analysis of the instrument's credibility on the date of calibration and relate only to the Instrument calibrated.*

Date of Issue: 30/08/2020



  
Approved Signatory

**Cyrix Healthcare Pvt. Ltd.,**  
30/641 B Petta Junction, Poonithura, Kochi-38.  
Ph : 97455 55073  
Email: sales.kl@cyrix.in  
service.kl@cyrix.in

**CYRIX**<sup>TM</sup>  
HEALTH CARE PVT LTD

[www.cyrixhealthcare.com](http://www.cyrixhealthcare.com)

ISO Certified Company

AERB Approved  
QA Agency

NABL Accredited  
Calibration Agency

# CERTIFICATE OF CALIBRATION

CUSTOMER NAME	DHARMAGIRI ST.JOSEPH'S HOSPITAL		
ADDRESS	Koyilandy - Edavanna Rd, Agastiamuzhi, Mukkam, Kerala 673602	DATE OF CALIBRATION	29/08/2020
CERTIFICATE NO:	CHPL/PC/SJH/009	VALID UP TO	28/02/2021
		DATE OF ISSUE	30/08/2020

INSTRUMENTS DETAILS			
INSTRUMENT NAME	MICRO PIPETTE	SI.NO:	HR62975
MAKE/MODEL.NO:	ERBA	VISUAL INSPECTION	OK


DETAILS OF REFERENCE STANDARDS & MAJOR EQUIPMENTS USED		
EQUIPMENTS NAME	ANALYTICAL BALANCE	THERMO HYGROMETER
MAKE	WENSAR	HTC-1/-
MODEL/SI.NO:	MAB182/39736	..
CALIBRATION VALIDITY	13/02/2021	23/09/2020
CALIBRATION BY	STIC	STIC

ENVIRONMENTAL CONDITION			
TEMPERATURE	25.6°C	RELATIVE HUMIDITY	40.1%


CALIBRATION RESULTS							
SL NO	SET READING ( $\mu$ l)	MEASURED READING ( $\mu$ l)					STATUS
		I	II	III	IV	V	
1	5	5.0	5.01	5.02	5.031	5.01	P
2	10	10.03	10.02	10.02	10.04	10.03	P
3	50	50.03	50.21	50.12	50.2	50.1	P

REMARKS: -

- ❖ This result of calibration refers to only to the particular items submitted for calibration.
- ❖ The above result is valid at the time of and under the stated conditions measurement.
- ❖ Calibration certificate issued for weight & measure parameters like mass, weighting balance, volumetric equipment, measuring scales/tapes etc. Are scientific purpose only and should not use for commercial.

  
Calibrated By  
(SE)

  
Approved By  
(QM)  
SARATH K B

  
**Cyrix Healthcare Pvt. Ltd.,**  
30/641 B Petta Junction, Poonithura, Kochi-38.  
Ph: 97455 55073  
Email: sales.kl@cyrix.in  
service.kl@cyrix.in

**CYRIX**<sup>TM</sup>  
HEALTH CARE PVT LTD



# CERTIFICATE OF CALIBRATION

Ref. No.: SJH010  
Applicant : DHARMAGIRI ST JOSEPH'S HOSPITAL.  
Device Calibrated : MICRO PIPETTE  
Make/Model: BIOPETTE™  
Sl.No: 04032534

Condition at receipt : Functional  
Calibrated at : Onsite  
Date of Calibration : 29/08/2020  
Calibration due date : 28/02/2021

*Certified that the above Instrument has been calibrated by trained technical personnel of this centre using a calibration system traceable to National Standards. The calibration details attached with the certificate are authentic quantitative analysis of the instrument's credibility on the date of calibration and relate only to the Instrument calibrated.*

30/08/2020

Date of Issue :

  
Approved Signatory

**Cyrix Healthcare Pvt. Ltd.,**  
30/641 B Petta Junction, Poonithura, Kochi-38.  
Ph : 97455 55073  
Email: sales.kl@cyrix.in  
service.kl@cyrix.in

**CYRIX™**  
HEALTH CARE PVT LTD

[www.cyrixhealthcare.com](http://www.cyrixhealthcare.com)

ISO Certified Company

AERB Approved  
QA Agency

NABL Accredited  
Calibration Agency

# CERTIFICATE OF CALIBRATION

CUSTOMER NAME	DHARMAGIRI ST.JOSEPH'S HOSPITAL		
ADDRESS	Koyilandy - Edavanna Rd, Agastiamuzhi, Mukkam, Kerala 673602	DATE OF CALIBRATION	29/08/2020
CERTIFICATE NO:	CHPL/PC/SJH/010	VALID UP TO	28/02/2021
		DATE OF ISSUE	30/08/2020

INSTRUMENTS DETAILS			
INSTRUMENT NAME	MICRO PIPETTE	SI.NO:	04032534
MAKE/MODEL.NO:	BIOPETTE	VISUAL INSPECTION	OK

DETAILS OF REFERENCE STANDARDS & MAJOR EQUIPMENTS USED		
EQUIPMENTS NAME	ANALYTICAL BALANCE	THERMO HYGROMETER
MAKE	WENSAR	HTC-1/-
MODEL/SI.NO:	MAB182/39736	..
CALIBRATION VALIDITY	13/02/2021	23/09/2020
CALIBRATION BY	STIC	STIC

ENVIRONMENTAL CONDITION			
TEMPERATURE	25.6°C	RELATIVE HUMIDITY	40.1%

CALIBRATION RESULTS							
SL NO	SET READING ( $\mu$ l)	MEASURED READING ( $\mu$ l)					STATUS
		I	II	III	IV	V	
1	1000	1000.03	1000.2	1000.02	1000.04	1000.3	P

REMARKS: -

- ❖ This result of calibration refers to only to the particular items submitted for calibration.
- ❖ The above result is valid at the time of and under the stated conditions measurement.
- ❖ Calibration certificate issued for weight & measure parameters like mass, weighting balance, volumetric equipment, measuring scales/tapes etc. Are scientific purpose only and should not use for commercial.

*Jh*  
Calibrated By  
(SE)

*Feb*  
Approved By  
(QM)  
SARATH K B



**Cyrix Healthcare Pvt. Ltd.,**  
30/641 B Petta Junction, Poonithura, Kochi-38.  
Ph: 97455 55073  
Email: sales.kl@cyrix.in  
service.kl@cyrix.in

**CYRIX**<sup>TM</sup>  
HEALTH CARE PVT LTD



# CERTIFICATE OF CALIBRATION

Ref. No.: SJH004  
Applicant : DHARMAGIRI ST JOSEPH'S HOSPITAL.  
Device Calibrated : MICRO PIPETTE  
Make/Model: EPPENDROF  
Sl.No: 118422

Condition at receipt : Functional  
Calibrated at : Onsite  
Date of Calibration : 29/08/2020  
Calibration due date : 28/02/2021

*Certified that the above Instrument has been calibrated by trained technical personnel of this centre using a calibration system traceable to National Standards. The calibration details attached with the certificate are authentic quantitative analysis of the instrument's credibility on the date of calibration and relate only to the Instrument calibrated.*

Date of Issue : 30/08/2020



  
Approved Signatory

**Cyrux Healthcare Pvt. Ltd.,**  
30/641 B Petta Junction, Poonithura, Kochi-38.  
Ph : 97455 55073  
Email: sales.kl@cyrix.in  
service.kl@cyrix.in

**CYRIX**<sup>TM</sup>  
HEALTH CARE PVT LTD

[www.cyruxhealthcare.com](http://www.cyruxhealthcare.com)

ISO Certified Company

AERB Approved  
QA Agency

NABL Accredited  
Calibration Agency

# CERTIFICATE OF CALIBRATION

CUSTOMER NAME	DHARMAGIRI ST.JOSEPH'S HOSPITAL		
ADDRESS	Koyilandy - Edavanna Rd, Agastiamuzhi, Mukkam, Kerala 673602	DATE OF CALIBRATION	29/08/2020
CERTIFICATE NO:	CHPL/PC/SJH/004	VALID UP TO	28/02/2021
		DATE OF ISSUE	30/08/2020

INSTRUMENTS DETAILS			
INSTRUMENT NAME	MICRO PIPETTE	SI.NO:	118422
MAKE/MODEL.NO:	EPPENDROF	VISUAL INSPECTION	OK


DETAILS OF REFERENCE STANDARDS & MAJOR EQUIPMENTS USED		
EQUIPMENTS NAME	ANALYTICAL BALANCE	THERMO HYGROMETER
MAKE	WENSAR	HTC-1/-
MODEL/SI.NO:	MAB182/39736	..
CALIBRATION VALIDITY	13/02/2021	23/09/2020
CALIBRATION BY	STIC	STIC

ENVIRONMENTAL CONDITION			
TEMPERATURE	25.6°C	RELATIVE HUMIDITY	40.1%

CALIBRATION RESULTS							
SL NO	SET READING ( $\mu$ l)	MEASURED READING ( $\mu$ l)					STATUS
		I	II	III	IV	V	
1	100	100.21	100.2	100.15	100.1	100.02	P

REMARKS: -

- ❖ This result of calibration refers to only to the particular items submitted for calibration.
- ❖ The above result is valid at the time of and under the stated conditions measurement.
- ❖ Calibration certificate issued for weight & measure parameters like mass, weighting balance, volumetric equipment, measuring scales/tapes etc. Are scientific purpose only and should not use for commercial.

  
Calibrated By  
(SE)

  
Approved By  
(QM)  
SARATH K B



**Cyrix Healthcare Pvt. Ltd.,**  
30/641 B Petta Junction, Poonithura, Kochi-38.  
Ph: 97455 55073  
Email: sales.kl@cyrix.in  
service.kl@cyrix.in

**CYRIX**<sup>TM</sup>  
HEALTH CARE PVT LTD



# CERTIFICATE OF CALIBRATION

Ref. No.: SJ H005  
Applicant : DHARMAGIRI ST JOSEPH'S HOSPITAL.  
Device Calibrated : MICRO PIPETTE  
Make/Model: EPPENDROF  
SL.No: 64944

Condition at receipt : Functional  
Calibrated at : Onsite  
Date of Calibration : 29/08/2020  
Calibration due date : 28/02/2021

*Certified that the above Instrument has been calibrated by trained technical personnel of this centre using a calibration system traceable to National Standards. The calibration details attached with the certificate are authentic quantitative analysis of the instrument's credibility on the date of calibration and relate only to the Instrument calibrated.*

Date of Issue : 30/08/2020



*Sob*  
Approved Signatory

**Cyrix Healthcare Pvt. Ltd.,**  
30/641 B Petta Junction, Poonithura, Kochi-38.  
Ph : 97455 55073  
Email: sales.kl@cyrix.in  
service.kl@cyrix.in

**CYRIX**<sup>TM</sup>  
HEALTH CARE PVT LTD

[www.cyrixhealthcare.com](http://www.cyrixhealthcare.com)

ISO Certified Company

AERB Approved  
QA Agency

NABL Accredited  
Calibration Agency

# CERTIFICATE OF CALIBRATION

CUSTOMER NAME	DHARMAGIRI ST.JOSEPH'S HOSPITAL		
ADDRESS	Koyilandy - Edavanna Rd., Agastiamuzhi, Mukkam, Kerala 673602	DATE OF CALIBRATION	29/08/2020
CERTIFICATE NO:	CHPL/PC/SJH/005	VALID UP TO	28/02/2021
		DATE OF ISSUE	30/08/2020

## INSTRUMENTS DETAILS

INSTRUMENT NAME	MICRO PIPETTE	SI.NO:	64944
MAKE/MODEL.NO:	EPPENDROF	VISUAL INSPECTION	OK

## DETAILS OF REFERENCE STANDARDS & MAJOR EQUIPMENTS USED

EQUIPMENTS NAME	ANALYTICAL BALANCE	THERMO HYGROMETER
MAKE	WENSAR	HTC-1/-
MODEL/SI.NO:	MAB182/39736	..
CALIBRATION VALIDITY	13/02/2021	23/09/2020
CALIBRATION BY	STIC	STIC

## ENVIRONMENTAL CONDITION

TEMPERATURE	25.6°C	RELATIVE HUMIDITY	40.1%
-------------	--------	-------------------	-------

## CALIBRATION RESULTS

SL NO	SET READING ( $\mu$ l)	MEASURED READING ( $\mu$ l)					STATUS
		I	II	III	IV	V	
1	1000	1000.1	1000.0	1001.0	1000.4	1000.8	P

### REMARKS: -

- ❖ This result of calibration refers to only to the particular items submitted for calibration.
- ❖ The above result is valid at the time of and under the stated conditions measurement.
- ❖ Calibration certificate issued for weight & measure parameters like mass, weighting balance, volumetric equipment, measuring scales/tapes etc. Are scientific purpose only and should not use for commercial.

Calibrated By

(SE)

Approved By

(QM)

SARATH K B



**Cyrix Healthcare Pvt. Ltd.,**

30/641 B Petta Junction, Poonithura, Kochi-38.

Ph: 97455 55073

Email: sales.kl@cyrix.in

service.kl@cyrix.in

**CYRIX**™  
HEALTHCARE PVT LTD

www.cyrixhealthcare.com



# CERTIFICATE OF CALIBRATION

Ref. No.: SJH006  
Applicant : DHARMAGIRI ST JOSEPH'S HOSPITAL.  
Device Calibrated : MICRO PIPETTE  
Make/Model: EPPENDROF  
Sl.No: 109973

Condition at  
receipt :

Functional


Calibrated at : Onsite

Date of Calibration : 29/08/2020

Calibration due date : 28/02/2021

*Certified that the above Instrument has been calibrated by trained technical personnel of this centre using a calibration system traceable to National Standards. The calibration details attached with the certificate are authentic quantitative analysis of the instrument's credibility on the date of calibration and relate only to the Instrument calibrated.*

Date of Issue : 30/08/2020

  
Approved Signatory



**Cyrix Healthcare Pvt. Ltd.,**  
30/641 B Petta Junction, Poonithura, Kochi-38.  
Ph : 97455 55073  
Email: sales.kl@cyrix.in  
service.kl@cyrix.in

**CYRIX**<sup>TM</sup>  
HEALTH CARE PVT LTD

[www.cyrixhealthcare.com](http://www.cyrixhealthcare.com)

ISO Certified Company

AERB Approved  
QA Agency

NABL Accredited  
Calibration Agency

# CERTIFICATE OF CALIBRATION

CUSTOMER NAME	DHARMAGIRI ST.JOSEPH'S HOSPITAL		
ADDRESS	Koyilandy - Edavanna Rd, Agastiamuzhi, Mukkam, Kerala 673602	DATE OF CALIBRATION	29/08/2020
CERTIFICATE NO:	CHPL/PC/SJH/006	VALID UP TO	28/02/2021
		DATE OF ISSUE	30/08/2020

### INSTRUMENTS DETAILS

INSTRUMENT NAME	MICRO PIPETTE	SI.NO:	109973
MAKE/MODEL.NO:	EPPENDROF	VISUAL INSPECTION	OK

### DETAILS OF REFERENCE STANDARDS & MAJOR EQUIPMENTS USED

EQUIPMENTS NAME	ANALYTICAL BALANCE	THERMO HYGROMETER
MAKE	WENSAR	HTC-1/-
MODEL/SI.NO:	MAB182/39736	..
CALIBRATION VALIDITY	13/02/2021	23/09/2020
CALIBRATION BY	STIC	STIC

### ENVIRONMENTAL CONDITION

TEMPERATURE	25.6°C	RELATIVE HUMIDITY	40.1%
-------------	--------	-------------------	-------

### CALIBRATION RESULTS

SL NO	SET READING ( $\mu$ l)	MEASURED READING ( $\mu$ l)					STATUS
		I	II	III	IV	V	
1	500	500.1	500.0	501.0	499.9	500.8	P

REMARKS: -

- ❖ This result of calibration refers to only to the particular items submitted for calibration.
- ❖ The above result is valid at the time of and under the stated conditions measurement.
- ❖ Calibration certificate issued for weight & measure parameters like mass, weighting balance, volumetric equipment, measuring scales/tapes etc. Are scientific purpose only and should not use for commercial.

Calibrated By

(SE)

Approved By

(QM)

SARATH K B



**Cyrix Healthcare Pvt. Ltd.,**  
30/641 B Petta Junction, Poonithura, Kochi-38.  
Ph: 97455 55073  
Email: sales.kl@cyrix.in  
service.kl@cyrix.in

**CYRIX**<sup>TM</sup>  
HEALTH CARE PVT LTD

[www.cyrixhealthcare.com](http://www.cyrixhealthcare.com)



# CERTIFICATE OF CALIBRATION

Ref. No.: SJH002  
Applicant : DHARMAGIRI ST JOSEPH'S HOSPITAL.  
Device Calibrated : MICRO PIPETTE  
Make/Model: MISPA I2  
Sl.No: YN4A009216

Condition at receipt : Functional  
Calibrated at : Onsite  
Date of Calibration : 29/08/2020  
Calibration due date : 28/02/2021

*Certified that the above Instrument has been calibrated by trained technical personnel of this centre using a calibration system traceable to National Standards. The calibration details attached with the certificate are authentic quantitative analysis of the instrument's credibility on the date of calibration and relate only to the Instrument calibrated.*

Date of Issue : 30/08/2020



  
Approved Signatory

**Cyrix Healthcare Pvt. Ltd.,**  
30/641 B Petta Junction, Poonithura, Kochi-38.  
Ph : 97455 55073  
Email: sales.kl@cyrix.in  
service.kl@cyrix.in

**CYRIX**<sup>TM</sup>  
HEALTH CARE PVT LTD

[www.cyrixhealthcare.com](http://www.cyrixhealthcare.com)

ISO Certified Company

AERB Approved  
QA Agency

NABL Accredited  
Calibration Agency

# CERTIFICATE OF CALIBRATION

CUSTOMER NAME	DHARMAGIRI ST.JOSEPH'S HOSPITAL		
ADDRESS	Koyilandy - Edavanna Rd, Agastiamuzhi, Mukkam, Kerala 673602	DATE OF CALIBRATION	29/08/2020
CERTIFICATE NO:	CHPL/PC/SJH/002	VALID UP TO	28/02/2021
		DATE OF ISSUE	30/08/2020

INSTRUMENTS DETAILS			
INSTRUMENT NAME	MICRO PIPETTE	SI.NO:	YN4A009216
MAKE/MODEL.NO:	MISPA I2	VISUAL INSPECTION	OK

DETAILS OF REFERENCE STANDARDS & MAJOR EQUIPMENTS USED			
EQUIPMENTS NAME	ANALYTICAL BALANCE	THERMO HYGROMETER	
MAKE	WENSAR	HTC-1/-	
MODEL/SI.NO:	MAB182/39736	..	
CALIBRATION VALIDITY	13/02/2021	23/09/2020	
CALIBRATION BY	STIC	STIC	

ENVIRONMENTAL CONDITION			
TEMPERATURE	25.6°C	RELATIVE HUMIDITY	40.1%

CALIBRATION RESULTS							
SL NO	SET READING ( $\mu$ l)	MEASURED READING ( $\mu$ l)					STATUS
		I	II	III	IV	V	
1	20	19.97	19.91	19.92	19.93	19.94	P
2	100	100.2	100.1	100.10	100.1	100.02	P
3	200	199.99	200.1	200.01	200.2	200.2	P

**REMARKS: -**

- ❖ This result of calibration refers to only to the particular items submitted for calibration.
- ❖ The above result is valid at the time of and under the stated conditions measurement.
- ❖ Calibration certificate issued for weight & measure parameters like mass, weighting balance, volumetric equipment, measuring scales/tapes etc. Are scientific purpose only and should not use for commercial.

Calibrated By  
(SE)



Approved By  
(QM)

SARATH K B

**Cyrix Healthcare Pvt. Ltd.,**  
30/641 B Petta Junction, Poonithura, Kochi-38.  
Ph: 97455 55073  
Email: sales.kl@cyrix.in  
service.kl@cyrix.in

**CYRIX**<sup>TM</sup>  
HEALTH CARE PVT LTD

[www.cyrixhealthcare.com](http://www.cyrixhealthcare.com)



# CERTIFICATE OF CALIBRATION

Ref. No.: SJH003  
Applicant : DHARMAGIRI ST JOSEPH'S HOSPITAL.  
Device Calibrated : MICRO PIPETTE,  
Make/Model: MISPA 12  
Sl.No: AP19821

Condition at receipt : Functional  
Calibrated at : Onsite  
Date of Calibration : 29/08/2020  
Calibration due date : 28/02/2021

*Certified that the above Instrument has been calibrated by trained technical personnel of this centre using a calibration system traceable to National Standards. The calibration details attached with the certificate are authentic quantitative analysis of the instrument's credibility on the date of calibration and relate only to the Instrument calibrated.*

Date of Issue : 30/08/2020



*[Signature]*  
Approved Signatory

**Cyrix Healthcare Pvt. Ltd.,**  
30/641 B Petta Junction, Poonithura, Kochi-38.  
Ph : 97455 55073  
Email: sales.kl@cyrix.in  
service.kl@cyrix.in

**CYRIX**<sup>TM</sup>  
HEALTH CARE PVT LTD

[www.cyrixhealthcare.com](http://www.cyrixhealthcare.com)

ISO Certified Company

AERB Approved  
QA Agency

NABL Accredited  
Calibration Agency

# CERTIFICATE OF CALIBRATION

CUSTOMER NAME	<b>DHARMAGIRI ST. JOSEPH'S HOSPITAL</b>		
ADDRESS	Koyilandy - Edavanna Rd, Agastiamuzhi, Mukkam, Kerala 673602	DATE OF CALIBRATION	29/08/2020
		VALID UP TO	28/02/2021
CERTIFICATE NO:	CHPL/PC/SJH/003	DATE OF ISSUE	30/08/2020

INSTRUMENTS DETAILS			
INSTRUMENT NAME	MICRO PIPETTE	SI.NO:	AP19821
MAKE/MODEL.NO:	MISPA I2	VISUAL INSPECTION	OK


DETAILS OF REFERENCE STANDARDS & MAJOR EQUIPMENTS USED			
EQUIPMENTS NAME	ANALYTICAL BALANCE	THERMO HYGROMETER	
MAKE	WENSAR	HTC-1/-	
MODEL/SI.NO:	MAB182/39736	..	
CALIBRATION VALIDITY	13/02/2021	23/09/2020	
CALIBRATION BY	STIC	STIC	

ENVIRONMENTAL CONDITION		
TEMPERATURE	25.6°C	RELATIVE HUMIDITY
		40.1%

CALIBRATION RESULTS							
SL NO	SET READING (µl)	MEASURED READING (µl)					STATUS
		I	II	III	IV	V	
1	100	100.1	100.1	99.9	99.9	100.1	P
2	500	500.2	501.2	500.0	501.1	501.0	P
3	1000	1000.1	1001.4	999.8	999.6	999.9	P

**REMARKS: -**

- ❖ This result of calibration refers to only to the particular items submitted for calibration.
- ❖ The above result is valid at the time of and under the stated conditions measurement.
- ❖ Calibration certificate issued for weight & measure parameters like mass, weighting balance, volumetric equipment, measuring scales/tapes etc. Are scientific purpose only and should not use for commercial.

  
**Calibrated By**  
 (SE)



  
**Approved By**  
 (QM)  
 SARATH K B



# CERTIFICATE OF CALIBRATION

Ref. No.: SJH001  
Applicant : DHARMAGIRI ST JOSEPH'S HOSPITAL.  
Device Calibrated : MICRO PIPETTE  
Make/Model: MISPA I2  
Sl.No: LW03959

Condition at receipt : Functional  
Calibrated at : Onsite  
Date of Calibration : 29/08/2020  
Calibration due date : 28/02/2021

*Certified that the above Instrument has been calibrated by trained technical personnel of this centre using a calibration system traceable to National Standards. The calibration details attached with the certificate are authentic quantitative analysis of the instrument's credibility on the date of calibration and relate only to the Instrument calibrated.*

Date of Issue : 30/08/2020

  
Approved Signatory



**Cyrix Healthcare Pvt. Ltd.,**  
30/641 B Petta Junction, Poonithura, Kochi-38.  
Ph : 97455 55073  
Email: sales.kl@cyrix.in  
service.kl@cyrix.in

**CYRIX**<sup>TM</sup>  
HEALTH CARE PVT LTD

[www.cyrixhealthcare.com](http://www.cyrixhealthcare.com)

ISO Certified Company

AERB Approved  
QA Agency

NABL Accredited  
Calibration Agency

# CERTIFICATE OF CALIBRATION

CUSTOMER NAME	DHARMAGIRI ST.JOSEPH'S HOSPITAL		
ADDRESS	Koyilandy - Edavanna Rd, Agastiamuzhi, Mukkam, Kerala 673602	DATE OF CALIBRATION	29/08/2020
CERTIFICATE NO:	CHPL/PC/SJH/001	VALID UP TO	28/02/2021
		DATE OF ISSUE	30/08/2020

INSTRUMENTS DETAILS			
INSTRUMENT NAME	MICRO PIPETTE	SI.NO:	LW03959
MAKE/MODEL.NO:	MISPA I2	VISUAL INSPECTION	OK

DETAILS OF REFERENCE STANDARDS & MAJOR EQUIPMENTS USED		
EQUIPMENTS NAME	ANALYTICAL BALANCE	THERMO HYGROMETER
MAKE	WENSAR	HTC-1/-
MODEL/SI.NO:	MAB182/39736	..
CALIBRATION VALIDITY	13/02/2021	23/09/2020
CALIBRATION BY	STIC	STIC

ENVIRONMENTAL CONDITION			
TEMPERATURE	25.6°C	RELATIVE HUMIDITY	40.1%

CALIBRATION RESULTS							
SL NO	SET READING ( $\mu$ l)	MEASURED READING ( $\mu$ l)					STATUS
		I	II	III	IV	V	
1	2	2.0	2.01	2.01	2.02	2.02	P
2	10	10.03	10.02	10.02	10.04	10.03	P
3	20	20.03	20.03	20.04	20.04	20.02	P

REMARKS: -

- ❖ This result of calibration refers to only to the particular items submitted for calibration.
- ❖ The above result is valid at the time of and under the stated conditions measurement.
- ❖ Calibration certificate issued for weight & measure parameters like mass, weighting balance, volumetric equipment, measuring scales/tapes etc. Are scientific purpose only and should not use for commercial.

Calibrated By  
(SE)



Approved By  
(QM) Sarath K B

**Cyrix Healthcare Pvt. Ltd.,**  
30/641 B Petta Junction, Poonithura, Kochi-38.  
Ph: 97455 55073  
Email: sales.kl@cyrix.in  
service.kl@cyrix.in

**CYRIX**<sup>TM</sup>  
HEALTH CARE PVT LTD