


CERTIFICATE OF CALIBRATION

Ref. No.: SJH007
Applicant : DHARMAGIRI ST JOSEPH'S HOSPITAL.
Device Calibrated : MICRO PIPETTE
Make/Model: EPPENDORF
Sl.No: 117661

Condition at receipt : Functional
Calibrated at : Onsite
Date of Calibration : 29/08/2020
Calibration due date : 28/02/2021

Certified that the above Instrument has been calibrated by trained technical personnel of this centre using a calibration system traceable to National Standards. The calibration details attached with the certificate are authentic quantitative analysis of the instrument's credibility on the date of calibration and relate only to the Instrument calibrated.

Date of Issue : 30/08/2020


Approved Signatory



Cyrix Healthcare Pvt. Ltd.,
30/641 B Petta Junction, Poonithura, Kochi-38.
Ph : 97455 55073
Email: sales.kl@cyrix.in
service.kl@cyrix.in

CYRIXTM
HEALTH CARE PVT LTD

www.cyrixhealthcare.com

ISO Certified Company

AERB Approved
QA Agency

NABL Accredited
Calibration Agency

CERTIFICATE OF CALIBRATION

CUSTOMER NAME	DHARMAGIRI ST.JOSEPH'S HOSPITAL		
ADDRESS	Koyilandy - Edavanna Rd, Agastiamuzhi, Mukkam, Kerala 673602	DATE OF CALIBRATION	29/08/2020
CERTIFICATE NO:	CHPL/PC/SJH/007	VALID UP TO	28/02/2021
		DATE OF ISSUE	30/08/2020

INSTRUMENTS DETAILS			
INSTRUMENT NAME	MICRO PIPETTE	SI.NO:	117661
MAKE/MODEL.NO:	EPPENDROF	VISUAL INSPECTION	OK

DETAILS OF REFERENCE STANDARDS & MAJOR EQUIPMENTS USED			
EQUIPMENTS NAME	ANALYTICAL BALANCE	THERMO HYGROMETER	
MAKE	WENSAR	HTC-1/-	
MODEL/SI.NO:	MAB182/39736	..	
CALIBRATION VALIDITY	13/02/2021	23/09/2020	
CALIBRATION BY	STIC	STIC	

ENVIRONMENTAL CONDITION			
TEMPERATURE	25.6°C	RELATIVE HUMIDITY	40.1%

CALIBRATION RESULTS							
SL NO	SET READING (μ l)	MEASURED READING (μ l)					STATUS
		I	II	III	IV	V	
1	200	200.0	200.01	200.01	200.02	200.02	P

REMARKS: -

- ❖ This result of calibration refers to only to the particular items submitted for calibration.
- ❖ The above result is valid at the time of and under the stated conditions measurement.
- ❖ Calibration certificate issued for weight & measure parameters like mass, weighting balance, volumetric equipment, measuring scales/tapes etc. Are scientific purpose only and should not use for commercial.

Calibrated By

(SE)

Approved By

(QM)

SARATH K B



Cyrix Healthcare Pvt. Ltd.,
30/641 B Petta Junction, Poonithura, Kochi-38.
Ph: 97455 55073
Email: sales.kl@cyrix.in
service.kl@cyrix.in

CYRIXTM
HEALTH CARE PVT LTD

www.cyrixhealthcare.com

CERTIFICATE OF CALIBRATION

Ref. No.: SJH008
Applicant : DHARMAGIRI ST JOSEPH'S HOSPITAL.
Device Calibrated : MICRO PIPETTE
Make/Model: EPPENDORF
Sl.No: 115937

Condition at receipt : Functional
Calibrated at : Onsite
Date of Calibration : 29/08/2020
Calibration due date : 28/02/2021

Certified that the above Instrument has been calibrated by trained technical personnel of this centre using a calibration system traceable to National Standards. The calibration details attached with the certificate are authentic quantitative analysis of the instrument's credibility on the date of calibration and relate only to the Instrument calibrated.

Date of Issue : 30/08/2020



[Signature]
Approved Signatory

Cyrix Healthcare Pvt. Ltd.,
30/641 B Petta Junction, Poonithura, Kochi-38.
Ph : 97455 55073
Email: sales.kl@cyrix.in
service.kl@cyrix.in

CYRIXTM
HEALTH CARE PVT LTD

www.cyrixhealthcare.com

CERTIFICATE OF CALIBRATION

CUSTOMER NAME	DHARMAGIRI ST.JOSEPH'S HOSPITAL		
ADDRESS	Koyilandy - Edavanna Rd, Agastiamuzhi, Mukkam, Kerala 673602	DATE OF CALIBRATION	29/08/2020
CERTIFICATE NO:	CHPL/PC/SJH/008	VALID UP TO	28/02/2021
		DATE OF ISSUE	30/08/2020

INSTRUMENTS DETAILS			
INSTRUMENT NAME	MICRO PIPETTE	SI.NO:	115937
MAKE/MODEL.NO:	EPPENDROF	VISUAL INSPECTION	OK


DETAILS OF REFERENCE STANDARDS & MAJOR EQUIPMENTS USED		
EQUIPMENTS NAME	ANALYTICAL BALANCE	THERMO HYGROMETER
MAKE	WENSAR	HTC-1/-
MODEL/SI.NO:	MAB182/39736	..
CALIBRATION VALIDITY	13/02/2021	23/09/2020
CALIBRATION BY	STIC	STIC


ENVIRONMENTAL CONDITION			
TEMPERATURE	25.6°C	RELATIVE HUMIDITY	40.1%

CALIBRATION RESULTS							
SL NO	SET READING (μ l)	MEASURED READING (μ l)					STATUS
		I	II	III	IV	V	
1	20	21.03	20.03	20.04	20.04	20.02	P

REMARKS: -

- ❖ This result of calibration refers to only to the particular items submitted for calibration.
- ❖ The above result is valid at the time of and under the stated conditions measurement.
- ❖ Calibration certificate issued for weight & measure parameters like mass, weighting balance, volumetric equipment, measuring scales/tapes etc. Are scientific purpose only and should not use for commercial.


Calibrated By
(SE)


Approved By
(QM)
SARATH K B



Cyrix Healthcare Pvt. Ltd.,
30/641 B Petta Junction, Poonithura, Kochi-38.
Ph: 97455 55073
Email: sales.kl@cyrix.in
service.kl@cyrix.in

CYRIXTM
HEALTH CARE PVT LTD

www.cyrixhealthcare.com

CERTIFICATE OF CALIBRATION

Ref. No.: SJH009
Applicant : DHARMAGIRI ST JOSEPH'S HOSPITAL.
Device Calibrated : MICRO PIPETTE
Make/Model: ERBA
Sl.No: HR62975

Condition at receipt : Functional
Calibrated at : Onsite
Date of Calibration : 29/08/2020
Calibration due date : 28/02/2021

Certified that the above Instrument has been calibrated by trained technical personnel of this centre using a calibration system traceable to National Standards. The calibration details attached with the certificate are authentic quantitative analysis of the instrument's credibility on the date of calibration and relate only to the Instrument calibrated.

Date of Issue: 30/08/2020




Approved Signatory

Cyrix Healthcare Pvt. Ltd.,
30/641 B Petta Junction, Poonithura, Kochi-38.
Ph : 97455 55073
Email: sales.kl@cyrix.in
service.kl@cyrix.in

CYRIXTM
HEALTH CARE PVT LTD

www.cyrixhealthcare.com

ISO Certified Company

AERB Approved
QA Agency

NABL Accredited
Calibration Agency

CERTIFICATE OF CALIBRATION

CUSTOMER NAME	DHARMAGIRI ST.JOSEPH'S HOSPITAL		
ADDRESS	Koyilandy - Edavanna Rd, Agastiamuzhi, Mukkam, Kerala 673602	DATE OF CALIBRATION	29/08/2020
CERTIFICATE NO:	CHPL/PC/SJH/009	VALID UP TO	28/02/2021
		DATE OF ISSUE	30/08/2020

INSTRUMENTS DETAILS			
INSTRUMENT NAME	MICRO PIPETTE	SI.NO:	HR62975
MAKE/MODEL.NO:	ERBA	VISUAL INSPECTION	OK

DETAILS OF REFERENCE STANDARDS & MAJOR EQUIPMENTS USED		
EQUIPMENTS NAME	ANALYTICAL BALANCE	THERMO HYGROMETER
MAKE	WENSAR	HTC-1/-
MODEL/SI.NO:	MAB182/39736	..
CALIBRATION VALIDITY	13/02/2021	23/09/2020
CALIBRATION BY	STIC	STIC

ENVIRONMENTAL CONDITION			
TEMPERATURE	25.6°C	RELATIVE HUMIDITY	40.1%


CALIBRATION RESULTS							
SL NO	SET READING (μ l)	MEASURED READING (μ l)					STATUS
		I	II	III	IV	V	
1	5	5.0	5.01	5.02	5.031	5.01	P
2	10	10.03	10.02	10.02	10.04	10.03	P
3	50	50.03	50.21	50.12	50.2	50.1	P

REMARKS: -

- ❖ This result of calibration refers to only to the particular items submitted for calibration.
- ❖ The above result is valid at the time of and under the stated conditions measurement.
- ❖ Calibration certificate issued for weight & measure parameters like mass, weighting balance, volumetric equipment, measuring scales/tapes etc. Are scientific purpose only and should not use for commercial.

Calibrated By
(SE)

Approved By
(QM)
SARATH K B


Cyrix Healthcare Pvt. Ltd.,
30/641 B Petta Junction, Poonithura, Kochi-38.
Ph: 97455 55073
Email: sales.kl@cyrix.in
service.kl@cyrix.in

CYRIXTM
HEALTH CARE PVT LTD

www.cyrixhealthcare.com

CERTIFICATE OF CALIBRATION

Ref. No.: SJH010
Applicant : DHARMAGIRI ST JOSEPH'S HOSPITAL.
Device Calibrated : MICRO PIPETTE
Make/Model: BIOPETTE™
Sl.No: 04032534

Condition at receipt : Functional
Calibrated at : Onsite
Date of Calibration : 29/08/2020
Calibration due date : 28/02/2021

Certified that the above Instrument has been calibrated by trained technical personnel of this centre using a calibration system traceable to National Standards. The calibration details attached with the certificate are authentic quantitative analysis of the instrument's credibility on the date of calibration and relate only to the Instrument calibrated.

30/08/2020

Date of Issue :


Approved Signatory

Cyrix Healthcare Pvt. Ltd.,
30/641 B Petta Junction, Poonithura, Kochi-38.
Ph : 97455 55073
Email: sales.kl@cyrix.in
service.kl@cyrix.in

CYRIX™
HEALTH CARE PVT LTD

www.cyrixhealthcare.com

ISO Certified Company

AERB Approved
QA Agency

NABL Accredited
Calibration Agency

CERTIFICATE OF CALIBRATION

CUSTOMER NAME	DHARMAGIRI ST.JOSEPH'S HOSPITAL		
ADDRESS	Koyilandy - Edavanna Rd, Agastiamuzhi, Mukkam, Kerala 673602	DATE OF CALIBRATION	29/08/2020
CERTIFICATE NO:	CHPL/PC/SJH/010	VALID UP TO	28/02/2021
		DATE OF ISSUE	30/08/2020

INSTRUMENTS DETAILS			
INSTRUMENT NAME	MICRO PIPETTE	SI.NO:	04032534
MAKE/MODEL.NO:	BIOPETTE	VISUAL INSPECTION	OK

DETAILS OF REFERENCE STANDARDS & MAJOR EQUIPMENTS USED		
EQUIPMENTS NAME	ANALYTICAL BALANCE	THERMO HYGROMETER
MAKE	WENSAR	HTC-1/-
MODEL/SI.NO:	MAB182/39736	..
CALIBRATION VALIDITY	13/02/2021	23/09/2020
CALIBRATION BY	STIC	STIC

ENVIRONMENTAL CONDITION			
TEMPERATURE	25.6°C	RELATIVE HUMIDITY	40.1%

CALIBRATION RESULTS							
SL NO	SET READING (μ l)	MEASURED READING (μ l)					STATUS
		I	II	III	IV	V	
1	1000	1000.03	1000.2	1000.02	1000.04	1000.3	P

REMARKS: -

- ❖ This result of calibration refers to only to the particular items submitted for calibration.
- ❖ The above result is valid at the time of and under the stated conditions measurement.
- ❖ Calibration certificate issued for weight & measure parameters like mass, weighting balance, volumetric equipment, measuring scales/tapes etc. Are scientific purpose only and should not use for commercial.

Jh
Calibrated By
(SE)

Feb
Approved By
(QM)
SARATH K B



Cyrix Healthcare Pvt. Ltd.,
30/641 B Petta Junction, Poonithura, Kochi-38.
Ph: 97455 55073
Email: sales.kl@cyrix.in
service.kl@cyrix.in

CYRIXTM
HEALTH CARE PVT LTD

CERTIFICATE OF CALIBRATION

Ref. No.: SJH004
Applicant : DHARMAGIRI ST JOSEPH'S HOSPITAL.
Device Calibrated : MICRO PIPETTE
Make/Model: EPPENDORF
Sl.No: 118422

Condition at receipt : Functional
Calibrated at : Onsite
Date of Calibration : 29/08/2020
Calibration due date : 28/02/2021

Certified that the above Instrument has been calibrated by trained technical personnel of this centre using a calibration system traceable to National Standards. The calibration details attached with the certificate are authentic quantitative analysis of the instrument's credibility on the date of calibration and relate only to the Instrument calibrated.

Date of Issue : 30/08/2020



[Signature]
Approved Signatory

Cyrux Healthcare Pvt. Ltd.,
30/641 B Petta Junction, Poonithura, Kochi-38.
Ph : 97455 55073
Email: sales.kl@cyrix.in
service.kl@cyrix.in

CYRUXTM
HEALTH CARE PVT LTD

www.cyruxhealthcare.com

ISO Certified Company

AERB Approved
QA Agency

NABL Accredited
Calibration Agency

CERTIFICATE OF CALIBRATION

CUSTOMER NAME	DHARMAGIRI ST.JOSEPH'S HOSPITAL		
ADDRESS	Koyilandy - Edavanna Rd, Agastiamuzhi, Mukkam, Kerala 673602	DATE OF CALIBRATION	29/08/2020
CERTIFICATE NO:	CHPL/PC/SJH/004	VALID UP TO	28/02/2021
		DATE OF ISSUE	30/08/2020

INSTRUMENTS DETAILS			
INSTRUMENT NAME	MICRO PIPETTE	SI.NO:	118422
MAKE/MODEL.NO:	EPPENDROF	VISUAL INSPECTION	OK

DETAILS OF REFERENCE STANDARDS & MAJOR EQUIPMENTS USED		
EQUIPMENTS NAME	ANALYTICAL BALANCE	THERMO HYGROMETER
MAKE	WENSAR	HTC-1/-
MODEL/SI.NO:	MAB182/39736	..
CALIBRATION VALIDITY	13/02/2021	23/09/2020
CALIBRATION BY	STIC	STIC

ENVIRONMENTAL CONDITION			
TEMPERATURE	25.6°C	RELATIVE HUMIDITY	40.1%

CALIBRATION RESULTS							
SL NO	SET READING (μ l)	MEASURED READING (μ l)					STATUS
		I	II	III	IV	V	
1	100	100.21	100.2	100.15	100.1	100.02	P

REMARKS: -

- ❖ This result of calibration refers to only to the particular items submitted for calibration.
- ❖ The above result is valid at the time of and under the stated conditions measurement.
- ❖ Calibration certificate issued for weight & measure parameters like mass, weighting balance, volumetric equipment, measuring scales/tapes etc. Are scientific purpose only and should not use for commercial.

Calibrated By
(SE)

Approved By
(QM)
SARATH K B



Cyrix Healthcare Pvt. Ltd.,
30/641 B Petta Junction, Poonithura, Kochi-38.
Ph: 97455 55073
Email: sales.kl@cyrix.in
service.kl@cyrix.in

CYRIXTM
HEALTH CARE PVT LTD

www.cyrixhealthcare.com

CERTIFICATE OF CALIBRATION

Ref. No.: SJ H005
Applicant : DHARMAGIRI ST JOSEPH'S HOSPITAL.
Device Calibrated : MICRO PIPETTE
Make/Model: EPPENDROF
SL.No: 64944

Condition at receipt : Functional
Calibrated at : Onsite
Date of Calibration : 29/08/2020
Calibration due date : 28/02/2021

Certified that the above Instrument has been calibrated by trained technical personnel of this centre using a calibration system traceable to National Standards. The calibration details attached with the certificate are authentic quantitative analysis of the instrument's credibility on the date of calibration and relate only to the Instrument calibrated.

Date of Issue : 30/08/2020



Sob
Approved Signatory

Cyrix Healthcare Pvt. Ltd.,
30/641 B Petta Junction, Poonithura, Kochi-38.
Ph : 97455 55073
Email: sales.kl@cyrix.in
service.kl@cyrix.in

CYRIXTM
HEALTH CARE PVT LTD

www.cyrixhealthcare.com

ISO Certified Company

AERB Approved
QA Agency

NABL Accredited
Calibration Agency

CERTIFICATE OF CALIBRATION

CUSTOMER NAME	DHARMAGIRI ST.JOSEPH'S HOSPITAL		
ADDRESS	Koyilandy - Edavanna Rd., Agastiamuzhi, Mukkam, Kerala 673602	DATE OF CALIBRATION	29/08/2020
CERTIFICATE NO:	CHPL/PC/SJH/005	VALID UP TO	28/02/2021
		DATE OF ISSUE	30/08/2020

INSTRUMENTS DETAILS

INSTRUMENT NAME	MICRO PIPETTE	SI.NO:	64944
MAKE/MODEL.NO:	EPPENDROF	VISUAL INSPECTION	OK

DETAILS OF REFERENCE STANDARDS & MAJOR EQUIPMENTS USED

EQUIPMENTS NAME	ANALYTICAL BALANCE	THERMO HYGROMETER
MAKE	WENSAR	HTC-1/-
MODEL/SI.NO:	MAB182/39736	..
CALIBRATION VALIDITY	13/02/2021	23/09/2020
CALIBRATION BY	STIC	STIC

ENVIRONMENTAL CONDITION

TEMPERATURE	25.6°C	RELATIVE HUMIDITY	40.1%
-------------	--------	-------------------	-------

CALIBRATION RESULTS

SL NO	SET READING (μ l)	MEASURED READING (μ l)					STATUS
		I	II	III	IV	V	
1	1000	1000.1	1000.0	1001.0	1000.4	1000.8	P

REMARKS: -

- ❖ This result of calibration refers to only to the particular items submitted for calibration.
- ❖ The above result is valid at the time of and under the stated conditions measurement.
- ❖ Calibration certificate issued for weight & measure parameters like mass, weighting balance, volumetric equipment, measuring scales/tapes etc. Are scientific purpose only and should not use for commercial.

Calibrated By

(SE)

Approved By

(QM)

SARATH K B



Cyrix Healthcare Pvt. Ltd.,
30/641 B Petta Junction, Poonithura, Kochi-38.
Ph: 97455 55073
Email: sales.kl@cyrix.in
service.kl@cyrix.in

CYRIX™
HEALTHCARE PVT LTD

www.cyrixhealthcare.com


CERTIFICATE OF CALIBRATION

Ref. No.: SJH006
Applicant : DHARMAGIRI ST JOSEPH'S HOSPITAL.
Device Calibrated : MICRO PIPETTE
Make/Model: EPPENDROF
Sl.No: 109973

Condition at receipt : Functional
Calibrated at : Onsite
Date of Calibration : 29/08/2020
Calibration due date : 28/02/2021

Certified that the above Instrument has been calibrated by trained technical personnel of this centre using a calibration system traceable to National Standards. The calibration details attached with the certificate are authentic quantitative analysis of the instrument's credibility on the date of calibration and relate only to the Instrument calibrated.

Date of Issue : 30/08/2020


Approved Signatory



Cyrix Healthcare Pvt. Ltd.,
30/641 B Petta Junction, Poonithura, Kochi-38.
Ph : 97455 55073
Email: sales.kl@cyrix.in
service.kl@cyrix.in

CYRIXTM
HEALTH CARE PVT LTD

www.cyrixhealthcare.com

ISO Certified Company

AERB Approved
QA Agency

NABL Accredited
Calibration Agency

CERTIFICATE OF CALIBRATION

CUSTOMER NAME	DHARMAGIRI ST.JOSEPH'S HOSPITAL		
ADDRESS	Koyilandy - Edavanna Rd, Agastiamuzhi, Mukkam, Kerala 673602	DATE OF CALIBRATION	29/08/2020
CERTIFICATE NO:	CHPL/PC/SJH/006	VALID UP TO	28/02/2021
		DATE OF ISSUE	30/08/2020

INSTRUMENTS DETAILS

INSTRUMENT NAME	MICRO PIPETTE	SI.NO:	109973
MAKE/MODEL.NO:	EPPENDROF	VISUAL INSPECTION	OK

DETAILS OF REFERENCE STANDARDS & MAJOR EQUIPMENTS USED

EQUIPMENTS NAME	ANALYTICAL BALANCE	THERMO HYGROMETER
MAKE	WENSAR	HTC-1/-
MODEL/SI.NO:	MAB182/39736	..
CALIBRATION VALIDITY	13/02/2021	23/09/2020
CALIBRATION BY	STIC	STIC

ENVIRONMENTAL CONDITION

TEMPERATURE	25.6°C	RELATIVE HUMIDITY	40.1%
-------------	--------	-------------------	-------

CALIBRATION RESULTS

SL NO	SET READING (μ l)	MEASURED READING (μ l)					STATUS
		I	II	III	IV	V	
1	500	500.1	500.0	501.0	499.9	500.8	P

REMARKS: -

- ❖ This result of calibration refers to only to the particular items submitted for calibration.
- ❖ The above result is valid at the time of and under the stated conditions measurement.
- ❖ Calibration certificate issued for weight & measure parameters like mass, weighting balance, volumetric equipment, measuring scales/tapes etc. Are scientific purpose only and should not use for commercial.

Calibrated By

(SE)

Approved By

(QM)

SARATH K B



Cyrix Healthcare Pvt. Ltd.,
30/641 B Petta Junction, Poonithura, Kochi-38.
Ph: 97455 55073
Email: sales.kl@cyrix.in
service.kl@cyrix.in

CYRIXTM
HEALTH CARE PVT LTD

www.cyrixhealthcare.com

CERTIFICATE OF CALIBRATION

Ref. No.: SJH002
Applicant : DHARMAGIRI ST JOSEPH'S HOSPITAL.
Device Calibrated : MICRO PIPETTE
Make/Model: MISPA I2
Sl.No: YN4A009216

Condition at receipt : Functional
Calibrated at : Onsite
Date of Calibration : 29/08/2020
Calibration due date : 28/02/2021

Certified that the above Instrument has been calibrated by trained technical personnel of this centre using a calibration system traceable to National Standards. The calibration details attached with the certificate are authentic quantitative analysis of the instrument's credibility on the date of calibration and relate only to the Instrument calibrated.

Date of Issue : 30/08/2020




Approved Signatory

Cyrix Healthcare Pvt. Ltd.,
30/641 B Petta Junction, Poonithura, Kochi-38.
Ph : 97455 55073
Email: sales.kl@cyrix.in
service.kl@cyrix.in

CYRIXTM
HEALTH CARE PVT LTD

www.cyrixhealthcare.com

ISO Certified Company

AERB Approved
QA Agency

NABL Accredited
Calibration Agency

CERTIFICATE OF CALIBRATION

CUSTOMER NAME	DHARMAGIRI ST.JOSEPH'S HOSPITAL		
ADDRESS	Koyilandy - Edavanna Rd, Agastiamuzhi, Mukkam, Kerala 673602	DATE OF CALIBRATION	29/08/2020
CERTIFICATE NO:	CHPL/PC/SJH/002	VALID UP TO	28/02/2021
		DATE OF ISSUE	30/08/2020

INSTRUMENTS DETAILS			
INSTRUMENT NAME	MICRO PIPETTE	SI.NO:	YN4A009216
MAKE/MODEL.NO:	MISPA I2	VISUAL INSPECTION	OK

DETAILS OF REFERENCE STANDARDS & MAJOR EQUIPMENTS USED			
EQUIPMENTS NAME	ANALYTICAL BALANCE	THERMO HYGROMETER	
MAKE	WENSAR	HTC-1/-	
MODEL/SI.NO:	MAB182/39736	..	
CALIBRATION VALIDITY	13/02/2021	23/09/2020	
CALIBRATION BY	STIC	STIC	

ENVIRONMENTAL CONDITION			
TEMPERATURE	25.6°C	RELATIVE HUMIDITY	40.1%

CALIBRATION RESULTS							
SL NO	SET READING (μ l)	MEASURED READING (μ l)					STATUS
		I	II	III	IV	V	
1	20	19.97	19.91	19.92	19.93	19.94	P
2	100	100.2	100.1	100.10	100.1	100.02	P
3	200	199.99	200.1	200.01	200.2	200.2	P

REMARKS: -

- ❖ This result of calibration refers to only to the particular items submitted for calibration.
- ❖ The above result is valid at the time of and under the stated conditions measurement.
- ❖ Calibration certificate issued for weight & measure parameters like mass, weighting balance, volumetric equipment, measuring scales/tapes etc. Are scientific purpose only and should not use for commercial.

Calibrated By
(SE)



Approved By
(QM)

SARATH K B

Cyrix Healthcare Pvt. Ltd.,
30/641 B Petta Junction, Poonithura, Kochi-38.
Ph: 97455 55073
Email: sales.kl@cyrix.in
service.kl@cyrix.in

CYRIXTM
HEALTH CARE PVT LTD

CERTIFICATE OF CALIBRATION

Ref. No.: SJH003
Applicant : DHARMAGIRI ST JOSEPH'S HOSPITAL.
Device Calibrated : MICRO PIPETTE,
Make/Model: MISPA 12
Sl.No: AP19821

Condition at receipt : Functional
Calibrated at : Onsite
Date of Calibration : 29/08/2020
Calibration due date : 28/02/2021

Certified that the above Instrument has been calibrated by trained technical personnel of this centre using a calibration system traceable to National Standards. The calibration details attached with the certificate are authentic quantitative analysis of the instrument's credibility on the date of calibration and relate only to the Instrument calibrated.

Date of Issue : 30/08/2020



[Signature]
Approved Signatory

Cyrix Healthcare Pvt. Ltd.,
30/641 B Petta Junction, Poonithura, Kochi-38.
Ph : 97455 55073
Email: sales.kl@cyrix.in
service.kl@cyrix.in

CYRIXTM
HEALTH CARE PVT LTD

www.cyrixhealthcare.com

ISO Certified Company

AERB Approved
QA Agency

NABL Accredited
Calibration Agency

CERTIFICATE OF CALIBRATION

CUSTOMER NAME	DHARMAGIRI ST. JOSEPH'S HOSPITAL		
ADDRESS	Koyilandy - Edavanna Rd, Agastiamuzhi, Mukkam, Kerala 673602	DATE OF CALIBRATION	29/08/2020
CERTIFICATE NO:	CHPL/PC/SJH/003	VALID UP TO	28/02/2021
		DATE OF ISSUE	30/08/2020

INSTRUMENTS DETAILS			
INSTRUMENT NAME	MICRO PIPETTE	SI.NO:	AP19821
MAKE/MODEL.NO:	MISPA I2	VISUAL INSPECTION	OK

DETAILS OF REFERENCE STANDARDS & MAJOR EQUIPMENTS USED			
EQUIPMENTS NAME	ANALYTICAL BALANCE	THERMO HYGROMETER	
MAKE	WENSAR	HTC-1/-	
MODEL/SI.NO:	MAB182/39736	..	
CALIBRATION VALIDITY	13/02/2021	23/09/2020	
CALIBRATION BY	STIC	STIC	

ENVIRONMENTAL CONDITION		
TEMPERATURE	25.6°C	RELATIVE HUMIDITY
		40.1%

CALIBRATION RESULTS							
SL NO	SET READING (µl)	MEASURED READING (µl)					STATUS
		I	II	III	IV	V	
1	100	100.1	100.1	99.9	99.9	100.1	P
2	500	500.2	501.2	500.0	501.1	501.0	P
3	1000	1000.1	1001.4	999.8	999.6	999.9	P

REMARKS: -

- ❖ This result of calibration refers to only to the particular items submitted for calibration.
- ❖ The above result is valid at the time of and under the stated conditions measurement.
- ❖ Calibration certificate issued for weight & measure parameters like mass, weighting balance, volumetric equipment, measuring scales/tapes etc. Are scientific purpose only and should not use for commercial.

Calibrated By
(SE)



Approved By
(QM)
SARATH K B

Cyrix Healthcare Pvt. Ltd.,
30/641 B Petta Junction, Poonithura, Kochi-38.
Ph: 97455 55073
Email: sales.kl@cyrix.in
service.kl@cyrix.in



CERTIFICATE OF CALIBRATION

Ref. No.: SJH001
Applicant : DHARMAGIRI ST JOSEPH'S HOSPITAL.
Device Calibrated : MICRO PIPETTE
Make/Model: MISPA I2
Sl.No: LW03959

Condition at receipt : Functional
Calibrated at : Onsite
Date of Calibration : 29/08/2020
Calibration due date : 28/02/2021

Certified that the above Instrument has been calibrated by trained technical personnel of this centre using a calibration system traceable to National Standards. The calibration details attached with the certificate are authentic quantitative analysis of the instrument's credibility on the date of calibration and relate only to the Instrument calibrated.

Date of Issue : 30/08/2020


Approved Signatory



Cyrix Healthcare Pvt. Ltd.,
30/641 B Petta Junction, Poonithura, Kochi-38.
Ph : 97455 55073
Email: sales.kl@cyrix.in
service.kl@cyrix.in

CYRIXTM
HEALTH CARE PVT LTD

www.cyrixhealthcare.com

ISO Certified Company

AERB Approved
QA Agency

NABL Accredited
Calibration Agency

CERTIFICATE OF CALIBRATION

CUSTOMER NAME	DHARMAGIRI ST.JOSEPH'S HOSPITAL		
ADDRESS	Koyilandy - Edavanna Rd, Agastiamuzhi, Mukkam, Kerala 673602	DATE OF CALIBRATION	29/08/2020
CERTIFICATE NO:	CHPL/PC/SJH/001	VALID UP TO	28/02/2021
		DATE OF ISSUE	30/08/2020

INSTRUMENTS DETAILS			
INSTRUMENT NAME	MICRO PIPETTE	SI.NO:	LW03959
MAKE/MODEL.NO:	MISPA I2	VISUAL INSPECTION	OK

DETAILS OF REFERENCE STANDARDS & MAJOR EQUIPMENTS USED		
EQUIPMENTS NAME	ANALYTICAL BALANCE	THERMO HYGROMETER
MAKE	WENSAR	HTC-1/-
MODEL/SI.NO:	MAB182/39736	..
CALIBRATION VALIDITY	13/02/2021	23/09/2020
CALIBRATION BY	STIC	STIC

ENVIRONMENTAL CONDITION			
TEMPERATURE	25.6°C	RELATIVE HUMIDITY	40.1%

CALIBRATION RESULTS							
SL NO	SET READING (μ l)	MEASURED READING (μ l)					STATUS
		I	II	III	IV	V	
1	2	2.0	2.01	2.01	2.02	2.02	P
2	10	10.03	10.02	10.02	10.04	10.03	P
3	20	20.03	20.03	20.04	20.04	20.02	P

REMARKS: -

- ❖ This result of calibration refers to only to the particular items submitted for calibration.
- ❖ The above result is valid at the time of and under the stated conditions measurement.
- ❖ Calibration certificate issued for weight & measure parameters like mass, weighting balance, volumetric equipment, measuring scales/tapes etc. Are scientific purpose only and should not use for commercial.

Calibrated By
(SE)



Approved By
(QM) Sarath K B
SARATH K B

Cyrix Healthcare Pvt. Ltd.,
30/641 B Petta Junction, Poonithura, Kochi-38.
Ph: 97455 55073
Email: sales.kl@cyrix.in
service.kl@cyrix.in

CYRIXTM
HEALTH CARE PVT LTD

CERTIFICATE OF CALIBRATION

Ref. No.: SJH 165
Applicant: DHARMAGIRI ST. JOSEPH'S HOSPITAL
Device Calibrated: DIGITAL THERMOMETER
Make/Model: .. REFRIGERATOR - SAMSUNG /
Sl. No: .. RA 23BD TS1 / XTL

Condition at receipt: Functional
Calibrated at : Onsite
Date of Calibration : 3/06/2020
Calibration due date : 2/06/2021

Certified that the above Instrument has been calibrated by trained technical personnel of this centre using a calibration system traceable to National Standards. The calibration details attached with the certificate are authentic quantitative analysis of the instrument's credibility on the date of calibration and relate only to the Instrument calibrated.

Date of Issue : 4/06/2020


Approved Signatory



Cyrix Healthcare Pvt. Ltd.,
30/641 B Petta Junction, Poonithura, Kochi-38.
Ph : 97455 55073
Email: sales.kl@cyrix.in
service.kl@cyrix.in

CYRIXTM
HEALTH CARE PVT LTD

www.cyrixhealthcare.com

ISO Certified Company

AERB Approved
QA Agency

NABL Accredited
Calibration Agency

CERTIFICATE OF CALIBRATION

Customer Name & Address	DHARMAGIRI ST. JOSEPH'S HOSPITAL, AGASTIAMUZZHI MUKKAM, KOZHICKODE KERALA - 673602	Certificate No.	SJH 165
		Date of Calibration	3/06/2020
		Valid Up to	2/06/2021
		Date of Issue	4/06/2020

Instruments Details			
Instrument Name	DIGITAL THERMOMETER	S No.	..
Make/Model No.	..	LOCATION	LABORATORY
BME ID	SJH/LAB/DTM/01	Visual Inspection	OK

Details of Reference Standards & Major Equipment's Used					
No	Equipment Name	Make/Model	Serial No:	Calibration Validity	Calibration By
1	INFRARED THERMOMETER	HTC	2017084051	13/01/2021	PERFECT CALIBRATION
2	DIGITAL THERMOHYGROMETER	-/HTC-2	..	23/09/2020	STIC

Environmental Condition			
Temperature	23.9°C	Relative Humidity	54.4 %

Calibration Results				
SL NO	SET READING(°C)	MEASURED READING (°C)	TOLERANCE	STATUS
1	2	1.6	±2	P
2	4	4.0	±2	P
3	6	6.0	±2	P

REMARKS: -


- ❖ This result of calibration refers to only to the particular items submitted for calibration.
- ❖ The above result is valid at the time of and under the stated conditions measurement.
- ❖ Calibration certificate issued for weight & measure parameters like mass, weighting balance, volumetric equipment, measuring scales/tapes etc. Are scientific purpose only and should not use for commercial.

Calibrated By
(SE)



Approved By
(QM)

SARATH K B



Cyrix Healthcare Pvt. Ltd.,
30/641 B Petta Junction, Poonithura, Kochi-38.
Ph: 97455 55073
Email: sales.kl@cyrix.in
service.kl@cyrix.in

CYRIX™
HEALTH CARE PVT LTD

CERTIFICATE OF CALIBRATION

Ref. No.: SJH 166
Applicant: DHARMAGIRI ST. JOSEPH'S HOSPITAL
Device Calibrated: DIGITAL THERMOMETER
Make/Model: .. REFRIGERATOR. SAMSUNG 10622
Sl. No: .. RA 23 BOTST/1XTL

Condition at receipt: Functional Calibrated at : Onsite
Date of Calibration : 3/06/2020
Calibration due date : 2/06/2021

Certified that the above Instrument has been calibrated by trained technical personnel of this centre using a calibration system traceable to National Standards. The calibration details attached with the certificate are authentic quantitative analysis of the instrument's credibility on the date of calibration and relate only to the Instrument calibrated.

Date of Issue :

4/06/2020

Approved Signatory



Cyrix Healthcare Pvt. Ltd.,
30/641 B Petta Junction, Poonithura, Kochi-38.
Ph : 97455 55073
Email: sales.kl@cyrix.in
service.kl@cyrix.in

CYRIXTM
HEALTH CARE PVT LTD

www.cyrixhealthcare.com

ISO Certified Company

AERB Approved
QA Agency

NABL Accredited
Calibration Agency

CERTIFICATE OF CALIBRATION

Customer Name & Address	DHARMAGIRI ST. JOSEPH'S HOSPITAL, AGASTIAMUZHI MUKKAM, KOZHIKKODE KERALA - 673602	Certificate No.	SJH 166
		Date of Calibration	3/06/2020
		Valid Up to	2/06/2021
		Date of Issue	4/06/2020

Instruments Details

Instrument Name	DIGITAL THERMOMETER	S No.	..
Make/Model No.	..	LOCATION	LABORATORY
BME ID	SJH/LAB/DTM/02	Visual Inspection	OK

Details of Reference Standards & Major Equipment's Used

No	Equipment Name	Make/Model	Serial No:	Calibration Validity	Calibration By
1	INFRARED THERMOMETER	HTC	2017084051	13/01/2021	PERFECT CALIBRATION
2	DIGITAL THERMOHYGROMETER	-/HTC-2	..	23/09/2020	STIC

Environmental Condition

Temperature	23.9°C	Relative Humidity	54.4 %
-------------	--------	-------------------	--------

Calibration Results

SL NO	SET READING(°C)	MEASURED READING (°C)	TOLERANCE	STATUS
1	2	1.6	±2	P
2	4	3.8	±2	P
3	6	6.0	±2	P

REMARKS: -

- ❖ This result of calibration refers to only to the particular items submitted for calibration.
- ❖ The above result is valid at the time of and under the stated conditions measurement.
- ❖ Calibration certificate issued for weight & measure parameters like mass, weighting balance, volumetric equipment, measuring scales/tapes etc. Are scientific purpose only and should not use for commercial.

Calibrated By
(SE)



Approved By
(QM)
SARATH K B

Cyrix Healthcare Pvt. Ltd.,
30/641 B Petta Junction, Poonithura, Kochi-38.
Ph: 97455 55073
Email: sales.kl@cyrix.in
service.kl@cyrix.in

CYRIX™
HEALTH CARE PVT LTD

CERTIFICATE OF CALIBRATION

Ref. No.: SJH 168
Applicant: DHARMAGIRI ST. JOSEPH'S HOSPITAL
Device Calibrated: CENTRIFUGE
Make/Model: ACCUVET
Sl. No: LD178AF001180

Condition at receipt: Functional
Calibrated at : Onsite
Date of Calibration : 3/06/2020
Calibration due date : 2/06/2021

Certified that the above Instrument has been calibrated by trained technical personnel of this centre using a calibration system traceable to National Standards. The calibration details attached with the certificate are authentic quantitative analysis of the instrument's credibility on the date of calibration and relate only to the Instrument calibrated.

Date of Issue : 4/06/2020


Approved Signatory



Cyrix Healthcare Pvt. Ltd.,
30/641 B Petta Junction, Poonithura, Kochi-38.
Ph : 97455 55073
Email: sales.kl@cyrix.in
service.kl@cyrix.in

CYRIXTM
HEALTH CARE PVT LTD

www.cyrixhealthcare.com

ISO Certified Company

AERB Approved
QA Agency

NABL Accredited
Calibration Agency

CERTIFICATE OF CALIBRATION

Customer Name & Address	DHARMAGIRI ST. JOSEPH'S HOSPITAL, AGASTIAMUZZHI MUKKAM, KOZHIKKODE KERALA - 673602	Certificate No.	SJH 168
		Date of Calibration	3/06/2020
		Valid Up to	2/06/2021
		Date of Issue	4/06/2020

Instruments Details

Instrument Name	CENTRIFUGE	S No.	LD178AF001180
Make/Model No.	ACCUVET	Location	LABORATORY
ID No.	SJH/LAB/CEN/01	Visual Inspection	OK

Details of Reference Standards & Major Equipment's Used

Equipment Name	DIGITAL TACHO METER	THERMO HYGROMETER
Make	METRAVI	HTC-2
Model/SR No.	NCTM 1000/07089699	...
Calibration Validity	16/01/2021	23/09/2020
Calibration By	STIC	STIC

Environmental Condition

Temperature	27°C	Relative Humidity	59%
-------------	------	-------------------	-----

Calibration Results

SL NO	SET READING (RPM)	MEASURED READING (RPM)	TOLERANCE(°C)	STATUS
1	1000	962	±5%	P
2	2000	1986	±5%	P
3	3000	2991	±5%	P

REMARKS: -

- ❖ This result of calibration refers to only to the particular items submitted for calibration.
- ❖ The above result is valid at the time of and under the stated conditions measurement.
- ❖ Calibration certificate issued for weight & measure parameters like mass, weighting balance, volumetric equipment, measuring scales/tapes etc. Are scientific purpose only and should not use for commercial.

Calibrated By

(SE)

Approved By

(QM)



Cyrix Healthcare Pvt. Ltd.,
30/641 B Petta Junction, Poonithura, Kochi-38.
Ph: 97455 55073
Email: sales.kl@cyrix.in
service.kl@cyrix.in

CYRIXTM
HEALTH CARE PVT LTD

www.cyrixhealthcare.com