

Authentic

Authentic Instrument &amp; Automation Pvt. Ltd.



Certificate No.: CC-2236

## CALIBRATION CERTIFICATE

ULR NO.: CC2236200000014853F  
Certificate No. 2319(I-5)/1003

Date of Calibration : 05.11.2020  
Next Suggested Date : 04.11.2021  
(As per agreed with the Customer)

Service Request Card No.	2319(I-5)	<b>ITEM'S DESCRIPTION &amp; IDENTIFICATION</b>	
Dated	03.11.2020	Nomenclature	MICRO PIPETTE
Condition of the Item	OK	Make	ERBA
<b>INDENTOR'S DETAILS</b>		Model	NA
M/s. DR. BOTHRAS DIAGNOSTIC AND IMAGING CENTER ,		Serial No.	PD548114
B 3-4 PAWAN PURI NEAR ,		ID No.	NA
RELIANCE FRESH BIKANER ,		Range	100-1000 ul
RAJASTHAN .		Least Count	1 ul
		Accuracy	NA
		Location	NA

## MAJOR EQUIPMENT USED

Nomenclature	ID No.	Traceability Certificate No.	Valid upto
Stainless Steel Laboratory Weights	WMCL/M/2019-02/2634	CC274319000000236F	27.02.2021
Weighing Balance (As a media)		as source	Nil

## ENVIRONMENTAL CONDITIONS

Temperature	: (20 ± 2)°C
Humidity	: (50 ± 10) %
Sound Level	: <60 dB
Illumination Level	: 450-700 Lux

## PROCEDURE FOLLOWED

MCP(L) - 20

## CALIBRATION RESULT OF.

## MICRO PIPETTE

*UUC Reading (µl)	Standard Reading (µl)	
	At 20°	At 27°
100	99.800	99.975
250	249.532	249.969
500	499.097	499.972
850	848.487	849.974
1000	998.221	999.971

Note : \*UUC : Unit Under Calibration i.e.  
The Expanded uncertainty associated with measurement at 95% confidence level  
and coverage factor k = 2 is ±0.8 µl

## MICRO PIPETTE

Checked By

Mr. Ravi Singh  
(Calibration Engineer)



Approved by

Mr. O. P. Singh  
(Technical Manager)

Address : 41, RICO Industrial Area, Near Kamani Choraha,  
Bopal, Jaipur - 302012 (Rajasthan) INDIA.  
Mobile : 099145-29160, 093524-54260, 099181-23798  
099181-23068

Email : calibration@authenticinstruments.com  
support@authenticinstruments.com  
qmt@authenticinstrument.com  
Website : www.authenticinstruments.com