



Certificate No.: CC-2236

CALIBRATION CERTIFICATE

ULR NO.: CC2236200000014856F

Certificate No. 2319(I-5)/1006

Date of Calibration : 05.11.2020

Next Suggested Date : 04.11.2021

(As per agreed with the Customer)

Service Request Card No.	2319(I-5)	ITEM'S DESCRIPTION & IDENTIFICATION	
Dated	03.11.2020	Nomenclature	Temp. Indicator
Condition of the Item	OK	Make	NA
INDENTOR'S DETAILS		Model	NA
M/s: DR. BOTHRAS DIAGNOSTIC AND IMAGING CENTER ,		Serial No.	NA
B 3-4 PAWAN PURI NEAR ,		ID No.	DBDC/DF/01
RELIANCE FRESH BIKANER ,		Range	(-10 TO 30 C)
RAJASTHAN .		Least Count	0.1 C
		Accuracy	NA
		Location	FREEZER

MAJOR EQUIPMENT USED

<i>Nomenclature</i>	<i>Ident. No.</i>	<i>Traceability Certificate No.</i>	<i>Valid upto</i>
RTD With Indicator	APLE-82	CC305219000000074F	25.09.2021

ENVIRONMENTAL CONDITIONS		PROCEDURE FOLLOWED
Temperature	: (25 ± 4)° C	TCP(L) - 03
Humidity	: (35 to 65) %	
Sound Level	: <60 dB	
Illumination Level	: 450-700 Lux	

CALIBRATION RESULT OF**Temp. Indicator**

Standard Reading	UUC Reading	Error	Uncertainty at 95 % C.L is ±
°C	°C	°C	(°C)
1.89	2.0	-0.11	0.97
3.84	4.0	-0.16	0.97
4.82	5.0	-0.18	0.97
5.92	6.0	-0.08	0.97
6.84	8.0	-1.16	0.97

Note : *UUC : Unit Under Calibration i.e.

Temp. Indicator

The Expanded uncertainty associated with measurement at 95% confidence level and coverage factor k = 2 is above under uncertainty column.