



Certificate No. CC-2236

CALIBRATION CERTIFICATE

ULR NO.: CC2236200000014856F
Certificate No. 2319(I-5)/1006

Date of Calibration : 05.11.2020
Next Suggested Date : 04.11.2021
(As per agreed with the Customer)

Service Request Card No.	2319(I-5)	ITEM'S DESCRIPTION & IDENTIFICATION	
Dated	03.11.2020	Nomenclature	Temp. Indicator
Condition of the Item	OK	Make	NA
INDENTOR'S DETAILS		Model	NA
M/s: DR. BOTHRAS DIAGNOSTIC AND IMAGING CENTER ,		Serial No.	NA
B 3-4 PAWAN PURI NEAR ,		ID No.	DBDC/DF/01
RELIANCE FRESH BIKANER ,		Range	(-10 TO 30 C)
RAJASTHAN .		Least Count	0.1 C
		Accuracy	NA
		Location	FREEZER

MAJOR EQUIPMENT USED

Nomenclature	Ident. No.	Traceability Certificate No.	Valid upto
RTD With Indicator	APLE-82	CC305219000000074F	25.09.2021

ENVIRONMENTAL CONDITIONS

Temperature	: (25 ± 4)° C
Humidity	: (35 to 65) %
Sound Level	: <60 dB
Illumination Level	: 450-700 Lux

PROCEDURE FOLLOWED

TCP(L) - 03

CALIBRATION RESULT OF**Temp. Indicator**

Standard Reading °C	UUC Reading °C	Error °C	Uncertainty at 95 % C.L is ± (°C)
1.89	2.0	-0.11	0.97
3.84	4.0	-0.16	0.97
4.82	5.0	-0.18	0.97
5.92	6.0	-0.08	0.97
6.84	8.0	-1.16	0.97

Note : *UUC : Under Calibration i.e.

Temp. Indicator

The Expanded uncertainty associated with measurement at 95% confidence level and coverage factor k = 2 is above under uncertainty column.

Checked By



Approved by

Mr. Ravi Singh
(Calibration Engineer)

Mr. O. P. Singh
(Technical Manager)

Address : M. PICO Industrial Area, Near Kamani Choraha,
Jodhpura, Jaipur - 302012 (Rajasthan) INDIA.
Mobile : 093145-29100, 093524-54200, 099181-23798
099181-23068

Email : calibration@authenticinstruments.com
support@authenticinstruments.com
qnt@authenticinstrument.com
Website : www.authenticinstruments.com