



UNMATCHED SERVICE

Date 30-10-2020

Calibration certificate

Customer Name : Mayom Hospital, Gurgaon

Instrument : EM200

Serial : B140505

Calibration Date : 30-10-2020

Calibration Due Date : 29-10-2021

This is to certified that we have calibrated Fully automated Biochemistry analyser Model EM200 Bearing serial no B140505 and on inspection of the instrument it is observed that result are well within the range and instrument working fine.

Instrument is properly calibrated

For Transasia Bio-Medicals Ltd

Dileep Kumar Yadav

Area Service manager

CALIBRATION CERTIFICATE

Prolyte Analyzer, Serial No. 30-170404383 installed at Mayom Hospital, D-Block, South City 1, Gurgram, Haryana-122001 is serviced on **06th Sep 2018** & calibrated on **06th Sep 2020**.

This instrument is calibrated through Calibration/ Quality Control Solution as given Below Detail:

Quality Control Kit – 92123

LOT NO. : 1801103

EXP.DT.: 2021-08

	Actual Value	Range
Normal		
	Na+ 136.3 mmol/L	129 - 144 mmol/L
	K+ 4.42 mmol/L	4.01 - 4.69 mmol/L
	Cl 94.8 mmol/L	87 - 103 mmol/L
Abnormal		
	Na+ 113.9 mmol/L	109 - 121 mmol/L
	K+ 2.08 mmol/L	1.86 – 2.18 mmol/L
	Cl 76.9 mmol/L	69 - 83 mmol/L

This is to certify that the above instrument is working satisfactory & Calibration is done for above Parameters.

The Next Scheduled Calibration is due on **06th Sep 2021**.

For **Aranax Medical Pvt. Ltd.**



Shubham Sharma
Service Engineer

Preventive Maintenance Schedule Chart

Name of Customer: - MAYOM HOSPITAL
D-Block, South City 1, Gurgram, Haryana-122001

Name of Instrument: - Electrolyte Analyzer- Prolyte (Na, K, Cl)

Serial No. of Instrument: - 30-170404383

Installation Date: - 06th Sep. 2018

Service Provider: - Aranax Medical Pvt.Ltd.
Office No- 312 A, Ansal Majestic Tower
Vikaspuri Delhi-110018

Name of Service Engineer: - Shubham Sharma
Mob: - +91-7217886041

Warranty Status: - In/Out of Warranty (tick any)

AMC Status: - In/Out of Warrant (tick any)

Preventive Maintenance Log: Year: - 2019-20

	Planned	Remark
P. Maintenance – 5	06 th September 2020	Done
P. Maintenance – 6	06 th March 2021	

For ARANAX MEDICAL PVT LTD.



Auth.Sign.

Aranax Medical Pvt. Ltd.

312 A, Ansal Majestic Tower, Vikas Puri, New Delhi-110018
+91 11 4942 4172 | info@aranaxmedical.com | www.aranaxmedical.com