


Calibration Certificate – Rx Imola

Instrument	Rx Imola Clinical Chemistry Analyser
Serial No	7201-0588
Account Name	APS Health Care, Guruvayoor
Installation Date	15 MARCH 2019
Calibration Date	07 SEP 2020
Next Calibration Due	07 MAR 2021

This is to certify that this analyser has been inspected and calibrated for following parameters:

Test Parameter	Target Value	Obtained Value
Input Voltage	230 -240 V AC	233 V AC
Cuvette Check	< 0.5 Abs for all cuvettes	< 0.11 Abs
Photometer Check	< 0.5 Abs for all wavelengths	< 0.14 Abs
Incubator Temperature	37° C ± 0.3 max	37.00 ° C
Reagent Tray Temperature	8-15°C	8.5°C
Detector Performance Check	8.5 to 9.5 Volts	8.6 to 9.4Volts
12V Lamp Supply	12 ± 0.3 volts	12.11 volts
5V Supply	5 ± 0.3 volts	5.03 volts
24V Supply	24 ± 0.3 volts	24.12 volts

The results obtained are as per specifications & tolerance ranges. The above calibration was done with a M/s. Vaidyanatheshwara (Model Mastech / MAS830L) digital Multimeter. (Calibration certificate is enclosed herewith).



Vipin. G
Customer Support Engineer