

**LAB MONTHLY SUMMARY**

Lab Name **DHARSHINI CLINICAL LAB AND E.C.G** Lab No **5929**  
 Month **October** Year **2020**  
 Constituent Group **Chemistry II**

Date of Result Entered : 19/10/2020

Date of Report Published : 09/11/2020

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE I	GOD-POD	Any Analyser	910	168.35	10.59	17.82	146 mg/dl	-1.25	1.18 ✓
2	UREA I	NED DYE	Any Analyser	98	110.31	10.53	11.62	102 mg/dl	-0.72	2.35 ✓
3	CREATININE I	JAFFE RATE-BLANKED & COMPENSATED/Alkaline Picrate	Any Analyser	456	4.50	11.88	0.54	4.3 mg/dl	-0.37	0.05 ✓
4	T.BILIRUBIN I	DIAZONIUM SALT (Colorimetric) /JENDRASSIK	Any Analyser	614	3.79	15.81	0.60	3.1 mg/dl	-1.15	0.05 ✓
5	T-PROTEIN I	BIURET - colorimetric	Any Analyser	667	5.06	10.46	0.53	5.5 g/dl	0.83	0.04 ✓
6	ALBUMIN I	BCG - colorimetric	Any Analyser	622	3.02	9.51	0.29	2.8 g/dl	-0.77	0.02 ✓
7	URIC ACID I	ENZYMATIC / URICASE Colorimetric	Any Analyser	517	4.49	18.56	0.83	5.1 mg/dl	0.73	0.07 ✓
8	CHOLESTEROL I	CHOD-PAP	Any Analyser	771	101.13	13.52	13.68	101 mg/dl	-0.01	0.98 ✓
9	TRIGLYCERIDE I	GPO-PAP / Enzymatic Colorimetric / End Point	Any Analyser	735	180.95	13.56	24.53	197 mg/dl	0.65	1.81 ✓

*Handwritten signature and date: 10/11/2020*

SDI Range	Interpretation
Within -1.0 to +1.0	Excellent.
Between ±1.0 to ±2.0	Good.
Between ±2.0 to ±3.0	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

Contact details:  
 Email: clnqc@cmcvellore.ac.in  
 Contact Number: 0416-2283102

*Pamela Christudoss*  
 Dr. Pamela Christudoss  
 CMC EQAS Co-Ordinator  
 Christian Medical College, Vellore

\*\*\*\*\* End of Report \*\*\*\*\*