

## LAB MONTHLY SUMMARY

Lab Name **NUCLEUS HEALTH CARE LABORATORY AND ECG** Lab No **13534**Month **October** Year **2020**Constituent Group **Chemistry II**Date of Result Entered : **19/10/2020**Date of Report Published : **09/11/2020**

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE I	DRY CHEMISTRY	FUJI DRYCHEM SERIES	17	164.22	5.04	8.28	155 mg/dl	-1.11	4.02
2	UREA I	DRY CHEMISTRY	FUJI DRYCHEM SERIES	14	120.02	3.59	4.30	117.2 mg/dl	-0.66	2.30
3	CREATININE I	DRY CHEMISTRY	FUJI DRYCHEM SERIES	19	5.40	6.83	0.37	5.3 mg/dl	-0.27	0.17
4	T.BILIRUBIN I	DRY CHEMISTRY	FUJI DRYCHEM SERIES	27	3.36	9.60	0.32	3.2 mg/dl	-0.50	0.12
5	T-PROTEIN I	DRY CHEMISTRY	FUJI DRYCHEM SERIES	21	4.69	5.22	0.24	4.6 g/dl	-0.37	0.11
6	ALBUMIN I	DRY CHEMISTRY	FUJI DRYCHEM SERIES	21	2.81	6.56	0.18	2.9 g/dl	0.49	0.08
7	URIC ACID I	DRY CHEMISTRY	FUJI DRYCHEM SERIES	25	3.84	4.06	0.16	3.8 mg/dl	-0.26	0.06
8	CHOLESTEROL I	DRY CHEMISTRY	FUJI DRYCHEM SERIES	18	92.10	11.46	10.56	80 mg/dl	-1.15	4.98
9	TRIGLYCERIDE I	DRY CHEMISTRY	FUJI DRYCHEM SERIES	23	200.33	11.56	23.15	195 mg/dl	-0.23	9.65

SDI Range	Interpretation
Within -1.0 to +1.0	Excellent.
Between $\pm 1.0$ to $\pm 2.0$	Good.
Between $\pm 2.0$ to $\pm 3.0$	Accept with caution. Warning Signal.
Beyond $\pm 3.0$	Unacceptable performance. Action Signal.

Page 1 of 1

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

Contact details:

Email: [clinqc@cmcvellore.ac.in](mailto:clinqc@cmcvellore.ac.in)

Contact Number: 0416-2283102



Dr. Pamela Christudoss  
 CMC EQAS Co-Ordinator  
 Christian Medical College, Vellore

\*\*\*\*\* End of Report \*\*\*\*\*