



# ASIAN TECHNOLOGY

(A HOUSE OF CALIBRATION)  
Cert. No. : CC-2239



## CALIBRATION CERTIFICATE

Customer name And Address	M/S: Shiv Diagnostics Maternity Centre. Near Spectrum Mall Coart Road Kashipur V S Nagar-244713,(U.P)	Certificate No.	CC223920600038153 F
		Service request No & date	P/15/11/11/2020
		Date of calibration	11/11/2020
		Valid up to	10/11/2021
		Date of issue	12/11/2020

Instrument Details			
Instrument name	Micro Pipette	Sr. No.	YE3K141238
Make / Model No.	Erba	Location	Lab
Range / Size	5 to 50 µl	Accuracy	----
Least Count	5 µl	Visual Inspection	OK
I.D. No.	MP-01		

Detail of reference standards & Major equipments used			
Equipment Name	A set of weight box	Digital Weighing Balance	
Make	Weigh India	AND	
Model / SR No.	----	GH-252	
Certificate No.	WMCL/E/2018-12/2533	CC205320000000016F	
Calibration Validity	26/12/2021	16/02/2021	
Calibration by	WEIGHTRONICS	Kelvin Research Centre	

Environmental Condition	Temperature	23± 3 °C	Calibration Reference	IS:1997
	Relative Humidity	50±20%	Calibration Procedure	CP-33

### Calibration Results

Serial No.	Nominal Capacity in (µl)	Measured Capacity in (µl)	Uncertainty At 95% C.L. (coverage factor k=2)
01.	10.0	10.02132	±0.10 µl
02.	20.0	19.98954	
03.	30.0	29.97515	
04.	40.0	39.96968	
05.	50.0	49.96274	

Remarks:

- ❖ This result of calibration refer to only to the particular item submitted for calibration.
- ❖ The above result are valid at the time of and under the stated conditions measurement.
- ❖ Calibration certificate issued for weight & measure parameters like mass, weighting balance, volumetric equipment, measuring scales/tapes etc. are scientific purpose only and should not used for commercial.

Calibrated By  
(Calibration Engg./TM)  
(MAHIPAL)  
Form No. 32



Approved By  
(QM/TM)  
(NEERAJ TYAGI)  
Page No. 1 of 1

----End of Report----