



ASIAN TECHNOLOGY

(A HOUSE OF CALIBRATION)
Cert. No. : CC-2239



CALIBRATION CERTIFICATE

Customer name And Address	M/S: Shiv Diagnostics Maternity Centre. Near Spectrum Mall Coart Road Kashipur V S Nagar-244713,(U.P)	Certificate No.	CC223920600038156 F
		Service request No & date	P/15/11/11/2020
		Date of calibration	11/11/2020
		Valid up to	10/11/2021
		Date of issue	12/11/2020

Instrument Details			
Instrument name	Centrifuge	Sr. No.	----
Make	Remi	Location	Lab
Range / Size	0 to 5000 rpm	Accuracy	-----
Least Count	10 rpm	Visual Inspection	OK
I.D. No.	LDC-11	Model No	C-85416

Detail of reference standards & Major equipments used			
Equipment Name	Digital Tachometer		
Make	Lutron		
Model / SR No.	AC92792		
Certificate No.	CC294019EL0000515F		
Calibration Validity	27/11/2020		
Calibration by	Digital Calibration Technology		

Environmental Condition	Temperature	20±2°C	Calibration Reference	IS: 12508
	Relative Humidity	50±10 %	Calibration Procedure	CP-37

Calibration Results			
Serial No.	Standard value in (rpm)	Measured value in (rpm)	Uncertainty At 95% C.L. (coverage factor K=2)
01.	Position: 1	501	±0.20%rdg
02.	Position: 2	1002	
03.	Position: 3	2003	
04.	Position: 4	3004	
05.	Position: 5	4955	

Remarks:

- ❖ This result of calibration refers to only to the particular item submitted for calibration.
- ❖ The above result is valid at the time of and under the stated conditions measurement.
- ❖ Calibration certificate issued for weight & measure parameters like mass, weighting balance, volumetric equipment, measuring scales/tapes etc. are scientific purpose only and should not used for commercial.

Calibrated By
(Calibration Engg./TM)
(MAHIPAL)
Form No.-32



Approved By
(QM/TM)
(NEERAJ YAGI)
Page No. 1 of 1

----End of Report----